

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024435.D
 Acq On : 29 Aug 2022 09:45
 Operator : SY/VA
 Sample : VW0829SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 29 22:54:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Aug 29 22:53:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	335076	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	329461	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	150527	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	175746	19.027	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.120%
7) Chloroethane-d5	2.873	69	154840	22.374	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.480%
11) 1,1-Dichloroethene-d2	4.007	63	142011	17.728	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.920%
21) 2-Butanone-d5	7.080	46	87223	51.269	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.540%
24) Chloroform-d	7.647	84	275195	25.546	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.200%
26) 1,2-Dichloroethane-d4	8.305	65	164432	25.170	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.680%
32) Benzene-d6	8.269	84	502820	26.637	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.560%
36) 1,2-Dichloropropane-d6	9.275	67	158373	26.978	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	107.920%
41) Toluene-d8	10.323	98	449551	25.048	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.200%
43) trans-1,3-Dichloroprop...	10.573	79	67676	24.634	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.520%
47) 2-Hexanone-d5	10.921	63	69401	52.828	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.660%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	153594	23.109	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	92.440%
66) 1,2-Dichlorobenzene-d4	13.847	152	168308	29.692	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	118.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024435.D
Acq On : 29 Aug 2022 09:45
Operator : SY/VA
Sample : VW0829SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 29 22:54:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Aug 29 22:53:57 2022
Response via : Initial Calibration

