

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024495.D
 Acq On : 30 Aug 2022 17:08
 Operator : SY/VA
 Sample : VW0829SBL04
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Aug 31 02:56:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Aug 31 02:55:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.836	114	316422	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	302441	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	134130	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	182090	20.876	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	83.520%	
7) Chloroethane-d5	2.873	69	136282	20.853	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	83.400%	
11) 1,1-Dichloroethene-d2	4.001	63	116642	15.419	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	61.680%	
21) 2-Butanone-d5	7.074	46	80594	50.165	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	100.340%	
24) Chloroform-d	7.647	84	231761	22.783	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	91.120%	
26) 1,2-Dichloroethane-d4	8.305	65	142323	23.070	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	92.280%	
32) Benzene-d6	8.269	84	415909	24.001	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	96.000%	
36) 1,2-Dichloropropane-d6	9.268	67	136437	25.317	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	101.280%	
41) Toluene-d8	10.323	98	359367	21.812	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	87.240%	
43) trans-1,3-Dichloroprop...	10.573	79	57141	22.658	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	90.640%	
47) 2-Hexanone-d5	10.921	63	62974	52.219	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	104.440%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	138808	22.750	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	91.000%	
66) 1,2-Dichlorobenzene-d4	13.853	152	137221	27.167	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	108.680%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024495.D
Acq On : 30 Aug 2022 17:08
Operator : SY/VA
Sample : VW0829SBL04
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Aug 31 02:56:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Aug 31 02:55:40 2022
Response via : Initial Calibration

