

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024458.D
 Acq On : 29 Aug 2022 20:31
 Operator : SY/VA
 Sample : N4362-07
 Misc : 5.71g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0FB8

Quant Time: Aug 30 00:25:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 30 00:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	420714	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	415824	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	186923	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	194376	16.760	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.040%
7) Chloroethane-d5	2.873	69	123891	14.258	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	57.040%
11) 1,1-Dichloroethene-d2	4.001	63	146918	14.607	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.440%
21) 2-Butanone-d5	7.080	46	64207	30.058	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	60.120%
24) Chloroform-d	7.647	84	228886	16.922	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	67.680%
26) 1,2-Dichloroethane-d4	8.305	65	133508	16.277	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.120%#
32) Benzene-d6	8.268	84	441734	18.541	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.160%
36) 1,2-Dichloropropane-d6	9.274	67	132856	17.931	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	71.720%
41) Toluene-d8	10.323	98	411215	18.153	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.600%
43) trans-1,3-Dichloroprop...	10.573	79	55171	15.911	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	63.640%
47) 2-Hexanone-d5	10.920	63	54987	33.163	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.320%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	120679	14.386	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	57.560%
66) 1,2-Dichlorobenzene-d4	13.847	152	134495	19.107	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.440%
Target Compounds						
					Qvalue	
34) Trichloroethene	9.085	95	5923	0.868	ug/L	91
46) Tetrachloroethene	10.859	164	1127250	212.714	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024458.D
Acq On : 29 Aug 2022 20:31
Operator : SY/VA
Sample : N4362-07
Misc : 5.71g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0FB8

Quant Time: Aug 30 00:25:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 30 00:24:06 2022
Response via : Initial Calibration

