




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024468.D
Acq On : 30 Aug 2022 00:37
Operator : SY/VA
Sample : N4345-07
Misc : 5.52g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBVS7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
Supervised By :Mahesh Dadoda 08/31/2022

Quant Time: Aug 30 06:27:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 30 00:24:06 2022
Response via : Initial Calibration



TIC: VW024468.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

3.934min (-0.079) 0.02 ug/L

response 177

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	70.50	75.14
65.00	22.70	22.60
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024468.D
 Acq On : 30 Aug 2022 00:37
 Operator : SY/VA
 Sample : N4345-07
 Misc : 5.52g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBVS7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/31/2022

Quant Time: Aug 30 06:27:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 30 00:24:06 2022
 Response via : Initial Calibration



TIC: VW024468.D\data.ms

(11) 1,1-Dichloroethene-d2 (S)

4.007min (-0.006) 16.84 ug/L m

response 127398

Ion	Exp%	Act%
63.00	100.00	100.00
98.00	70.50	0.10#
65.00	22.70	0.03#
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024468.D
 Acq On : 30 Aug 2022 00:37
 Operator : SY/VA
 Sample : N4345-07
 Misc : 5.52g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBVS7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/31/2022

Quant Time: Aug 30 06:27:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 30 00:24:06 2022
 Response via : Initial Calibration



TIC: VW024468.D\data.ms

(13) Acetone (T)

4.123min (+ 0.006) 7.95 ug/L m

response 7628

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	29.20	16.07
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
 Data File : VW024468.D
 Acq On : 30 Aug 2022 00:37
 Operator : SY/VA
 Sample : N4345-07
 Misc : 5.52g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DBVS7

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 08/30/2022
 Supervised By :Mahesh Dadoda 08/31/2022

Quant Time: Aug 30 06:27:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Aug 30 00:24:06 2022
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.842	114		316476	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117		301615	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152		130987	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.343	65		175045	20.065	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	80.240%		
7) Chloroethane-d5	2.873	69		113147	17.310	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	69.240%		
11) 1,1-Dichloroethene-d2	4.007	63		127398m	16.838	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	67.360%		
21) 2-Butanone-d5	7.080	46		62719	39.032	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	78.060%		
24) Chloroform-d	7.647	84		190744	18.747	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	75.000%		
26) 1,2-Dichloroethane-d4	8.305	65		112724	18.269	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	73.080%		
32) Benzene-d6	8.269	84		366425	21.204	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	84.800%		
36) 1,2-Dichloropropane-d6	9.274	67		108265	20.145	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	80.560%		
41) Toluene-d8	10.323	98		341966	20.813	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	83.240%		
43) trans-1,3-Dichloroprop...	10.579	79		43239	17.192	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery =	68.760%		
47) 2-Hexanone-d5	10.921	63		51815	43.083	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	86.160%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84		112140	18.430	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	73.720%		
66) 1,2-Dichlorobenzene-d4	13.853	152		102539	20.788	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	83.160%		
Target Compounds							
						Qvalue	
13) Acetone	4.123	43		7628m	7.955	ug/L	
16) Methylene chloride	4.916	84		22609	4.260	ug/L	80
34) Trichloroethene	9.092	95		3194	0.645	ug/L	86
46) Tetrachloroethene	10.860	164		126221	32.837	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082922\
Data File : VW024468.D
Acq On : 30 Aug 2022 00:37
Operator : SY/VA
Sample : N4345-07
Misc : 5.52g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DBVS7

Manual IntegrationsAPPROVED

Quant Time: Aug 30 06:27:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Aug 30 00:24:06 2022
Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 08/30/2022
Supervised By :Mahesh Dadoda 08/31/2022

