

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082924\
 Data File : VW030097.D
 Acq On : 29 Aug 2024 14:09
 Operator : SY/MD
 Sample : P3696-06
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Aug 29 21:33:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Aug 29 21:32:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	162714	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	156448	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	67727	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	55479	22.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.640%
7) Chloroethane-d5	2.905	69	41817	24.333	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.320%
11) 1,1-Dichloroethene-d2	4.021	65	22804	22.043	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	88.160%
21) 2-Butanone-d5	7.075	46	26217	54.315	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.640%
24) Chloroform-d	7.648	84	108028	24.619	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.480%
26) 1,2-Dichloroethane-d4	8.307	65	59215	26.577	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.320%
32) Benzene-d6	8.276	84	210430	22.787	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.160%
36) 1,2-Dichloropropane-d6	9.276	67	68898	25.465	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.880%
41) Toluene-d8	10.325	98	187942	22.360	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.440%
43) trans-1,3-Dichloroprop...	10.575	79	25441	22.859	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.440%
47) 2-Hexanone-d5	10.922	63	20972	50.647	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.300%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	52056	26.132	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	104.520%
66) 1,2-Dichlorobenzene-d4	13.848	152	61180	24.974	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.880%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW082924\
Data File : VW030097.D
Acq On : 29 Aug 2024 14:09
Operator : SY/MD
Sample : P3696-06
Misc : 5.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Aug 29 21:33:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM082624SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Aug 29 21:32:12 2024
Response via : Initial Calibration

