

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090219\
 Data File : VW012500.D
 Acq On : 02 Sep 2019 15:45
 Operator : SY/VA
 Sample : K4528-18
 Misc : 4.87G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-S4-(1.9)

Quant Time: Sep 03 08:13:58 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	201058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	324446	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	235813	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	69117	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	127508	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
35) Dibromofluoromethane	7.88	113	98497	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
50) Toluene-d8	10.32	98	378179	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
62) 4-Bromofluorobenzene	12.62	95	117890	42.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.18%	
Target Compounds						
20) Methylene Chloride	4.92	84	15450	6.689	ug/l	97
29) Tetrahydrofuran	7.54	42	3087	6.198	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090219\
Data File : VW012500.D
Acq On : 02 Sep 2019 15:45
Operator : SY/VA
Sample : K4528-18
Misc : 4.87G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-S4-(1.9)

Quant Time: Sep 03 08:13:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

