

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW090219\
 Data File : VW012504.D
 Acq On : 02 Sep 2019 17:28
 Operator : SY/VA
 Sample : K4529-02
 Misc : 2.97G/5ML/MSVOA W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 T9-12-13-T3-(0)

Quant Time: Sep 03 08:20:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
 Quant Title : SW846 8260
 QLast Update : Mon Aug 26 07:18:31 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	208789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	339943	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	260053	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	83677	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.31	65	128499	51.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.54%	
35) Dibromofluoromethane	7.88	113	101705	50.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.24%	
50) Toluene-d8	10.32	98	399183	48.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.04%	
62) 4-Bromofluorobenzene	12.62	95	114650	39.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.06%	
Target Compounds						
20) Methylene Chloride	4.92	84	21374	8.912	ug/l	96
29) Tetrahydrofuran	7.54	42	6337	12.251	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW090219\
Data File : VW012504.D
Acq On : 02 Sep 2019 17:28
Operator : SY/VA
Sample : K4529-02
Misc : 2.97G/5ML/MSVOA W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
T9-12-13-T3-(0)

Quant Time: Sep 03 08:20:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W082319S.M
Quant Title : SW846 8260
QLast Update : Mon Aug 26 07:18:31 2019
Response via : Initial Calibration

