

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090222\
 Data File : VW024541.D
 Acq On : 02 Sep 2022 18:21
 Operator : SY/VA
 Sample : N4483-12
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW360

Quant Time: Sep 03 00:06:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 02 23:58:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	292306	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	250301	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	73202	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.337	65	113145	19.394	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	77.560%	
7) Chloroethane-d5	2.873	69	102243	23.538	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	94.160%	
11) 1,1-Dichloroethene-d2	4.001	63	65382	9.518	ug/L	-0.01
Spiked Amount	25.000	Range 45 - 110	Recovery	=	38.080%#	
21) 2-Butanone-d5	7.080	46	67684	73.570	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	147.140%#	
24) Chloroform-d	7.647	84	202423	26.007	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	104.040%	
26) 1,2-Dichloroethane-d4	8.299	65	136351	32.438	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	129.760%	
32) Benzene-d6	8.269	84	289888	22.202	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	88.800%	
36) 1,2-Dichloropropane-d6	9.268	67	115603	30.259	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	121.040%#	
41) Toluene-d8	10.317	98	208901	16.800	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	67.200%	
43) trans-1,3-Dichloroprop...	10.579	79	37362	22.417	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	89.680%	
47) 2-Hexanone-d5	10.927	63	45957	70.226	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	140.460%#	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	131925	37.983	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	151.920%#	
66) 1,2-Dichlorobenzene-d4	13.847	152	71321	29.783	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	119.120%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090222\
Data File : VW024541.D
Acq On : 02 Sep 2022 18:21
Operator : SY/VA
Sample : N4483-12
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW360

Quant Time: Sep 03 00:06:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 02 23:58:31 2022
Response via : Initial Calibration

