

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090222\
 Data File : VW024542.D
 Acq On : 02 Sep 2022 18:46
 Operator : SY/VA
 Sample : N4483-13
 Misc : 4.46g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW361

Quant Time: Sep 03 00:06:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 02 23:58:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	348526	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	335802	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	142196	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	126456	18.179	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.720%
7) Chloroethane-d5	2.873	69	113629	21.940	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.760%
11) 1,1-Dichloroethene-d2	4.007	63	76903	9.389	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	37.560%#
21) 2-Butanone-d5	7.074	46	83100	75.756	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	151.520%#
24) Chloroform-d	7.641	84	235328	25.357	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	101.440%
26) 1,2-Dichloroethane-d4	8.305	65	143234	28.579	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	114.320%
32) Benzene-d6	8.269	84	356377	20.345	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.360%
36) 1,2-Dichloropropane-d6	9.268	67	134816	26.303	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.200%
41) Toluene-d8	10.323	98	284890	17.078	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	68.320%
43) trans-1,3-Dichloroprop...	10.579	79	52554	23.504	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.000%
47) 2-Hexanone-d5	10.920	63	64112	73.024	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	146.040%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	153709	32.987	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	131.960%#
66) 1,2-Dichlorobenzene-d4	13.853	152	141199	30.354	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	121.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090222\
Data File : VW024542.D
Acq On : 02 Sep 2022 18:46
Operator : SY/VA
Sample : N4483-13
Misc : 4.46g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW361

Quant Time: Sep 03 00:06:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 02 23:58:31 2022
Response via : Initial Calibration

