


Data Path : Z:\voasrv\HPCHEM1\MSVOA W\Data\VW090719\
 Data File : VW012716.D
 Acq On : 07 Sep 2019 13:09
 Operator : SY/VA
 Sample : BFB
 Misc : 5.00G/5ML/MSVOA W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W090719S.M
 Title : SW846 8260
 Last Update : Sun Sep 08 05:04:40 2019



AutoFind: Scans 1756, 1757, 1758; Background Corrected with Scan 1748

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	30480	PASS
75	95	30	60	55.1	74805	PASS
95	95	100	100	100.0	135834	PASS
96	95	5	9	6.5	8858	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	77.4	105154	PASS
175	174	5	9	6.4	6773	PASS
176	174	95	101	97.0	101986	PASS
177	176	5	9	7.2	7325	PASS