

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
 Data File : VW027000.D
 Acq On : 07 Sep 2023 13:24
 Operator : SY/MD
 Sample : 04330-05
 Misc : 6.51g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EZYX6

Quant Time: Sep 08 01:34:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 08 01:32:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	811199	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	686861	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	236186	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	199671	13.777	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	55.120%	
7) Chloroethane-d5	2.893	69	190970	16.327	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	65.320%	
11) 1,1-Dichloroethene-d2	4.027	65	74533	13.955	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	55.800%	
21) 2-Butanone-d5	7.075	46	142072	50.877	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	101.760%	
24) Chloroform-d	7.648	84	384568	15.932	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	63.720%	
26) 1,2-Dichloroethane-d4	8.307	65	249649	19.677	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	78.720%	
32) Benzene-d6	8.276	84	878322	19.495	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	78.000%	
36) 1,2-Dichloropropane-d6	9.270	67	293169	22.166	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	88.680%	
41) Toluene-d8	10.319	98	673730	16.517	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	66.080%	
43) trans-1,3-Dichloroprop...	10.575	79	100554	19.591	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	78.360%	
47) 2-Hexanone-d5	10.922	63	77710	62.933	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	125.860%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	171255	16.685	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	66.720%	
66) 1,2-Dichlorobenzene-d4	13.848	152	175397	21.472	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	85.880%	
Target Compounds					Qvalue	
13) Acetone	4.125	43	20953	8.569	ug/L	91
14) Carbon disulfide	4.387	76	47176	1.283	ug/L	99
25) Chloroform	7.673	83	70820	3.054	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
Data File : VW027000.D
Acq On : 07 Sep 2023 13:24
Operator : SY/MD
Sample : 04330-05
Misc : 6.51g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EZYX6

Quant Time: Sep 08 01:34:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 08 01:32:52 2023
Response via : Initial Calibration

