




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
 Data File : VW027008.D
 Acq On : 07 Sep 2023 16:37
 Operator : SY/MD
 Sample : 04332-01
 Misc : 3.83g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB9

Manual IntegrationsAPPROVED

Quant Time: Sep 08 01:36:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 08 01:32:52 2023
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 09/08/2023
 Supervised By :Mahesh Dadoda 09/08/2023



TIC: VW027008.D\data.ms

(13) Acetone (T)

4.125min (+ 0.006) 1.78 ug/L

response 5802

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	33.20	53.90
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
 Data File : VW027008.D
 Acq On : 07 Sep 2023 16:37
 Operator : SY/MD
 Sample : 04332-01
 Misc : 3.83g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB9

Manual IntegrationsAPPROVED

Quant Time: Sep 08 01:36:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 08 01:32:52 2023
 Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 09/08/2023
 Supervised By :Mahesh Dadoda 09/08/2023



TIC: VW027008.D\data.ms

(13) Acetone (T)

4.125min (+ 0.006) 3.38 ug/L m

response 10983

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	33.20	28.47
0.00	0.00	0.00
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
 Data File : VW027008.D
 Acq On : 07 Sep 2023 16:37
 Operator : SY/MD
 Sample : 04332-01
 Misc : 3.83g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AB9

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 09/08/2023
 Supervised By :Mahesh Dadoda 09/08/2023

Quant Time: Sep 08 01:36:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 08 01:32:52 2023
 Response via : Initial Calibration

Compound	R.T.	Q	Ion	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene	8.843	114		1078506	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117		874494	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152		321033	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3	2.369	65		231919	12.036	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	48.160%		
7) Chloroethane-d5	2.899	69		218812	14.071	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150		Recovery =	56.280%		
11) 1,1-Dichloroethene-d2	4.021	65		88354	12.442	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110		Recovery =	49.760%		
21) 2-Butanone-d5	7.081	46		71704	19.313	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	38.620%		
24) Chloroform-d	7.648	84		509535	15.877	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150		Recovery =	63.520%		
26) 1,2-Dichloroethane-d4	8.307	65		272188	16.136	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130		Recovery =	64.560%#		
32) Benzene-d6	8.276	84		1004039	17.504	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135		Recovery =	70.000%		
36) 1,2-Dichloropropane-d6	9.276	67		321107	19.069	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120		Recovery =	76.280%		
41) Toluene-d8	10.325	98		803131	15.465	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130		Recovery =	61.880%		
43) trans-1,3-Dichloroprop...	10.575	79		71822	10.991	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135		Recovery =	43.960%		
47) 2-Hexanone-d5	10.922	63		31663	20.140	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135		Recovery =	40.280%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84		254198	19.452	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120		Recovery =	77.800%		
66) 1,2-Dichlorobenzene-d4	13.849	152		186035	16.755	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120		Recovery =	67.000%#		
Target Compounds							
13) Acetone	4.125	43		10983m	3.379	ug/L	
16) Methylene chloride	4.936	84		47846	2.290	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090723\
Data File : VW027008.D
Acq On : 07 Sep 2023 16:37
Operator : SY/MD
Sample : 04332-01
Misc : 3.83g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AB9

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 09/08/2023
Supervised By :Mahesh Dadoda 09/08/2023

Quant Time: Sep 08 01:36:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090523SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 08 01:32:52 2023
Response via : Initial Calibration

