

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090922\
 Data File : VW024595.D
 Acq On : 09 Sep 2022 11:53
 Operator : SY/VA
 Sample : N4440-06
 Misc : 5.05g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Sep 10 00:40:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Sep 10 00:38:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	278141	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	280540	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	117468	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	148279	25.821	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.280%
7) Chloroethane-d5	2.873	69	113348	26.546	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	106.200%
11) 1,1-Dichloroethene-d2	4.001	63	90649	14.962	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.840%
21) 2-Butanone-d5	7.080	46	40498	43.717	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.440%
24) Chloroform-d	7.647	84	148071	20.902	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	83.600%
26) 1,2-Dichloroethane-d4	8.305	65	93164	23.326	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.320%
32) Benzene-d6	8.268	84	298259	21.614	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.440%
36) 1,2-Dichloropropane-d6	9.274	67	89030	21.882	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	87.520%
41) Toluene-d8	10.323	98	286567	21.302	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.200%
43) trans-1,3-Dichloroprop...	10.579	79	36506	20.022	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.080%
47) 2-Hexanone-d5	10.926	63	33298	44.524	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.040%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	78603	20.156	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	80.640%
66) 1,2-Dichlorobenzene-d4	13.853	152	87207	23.439	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW090922\
Data File : VW024595.D
Acq On : 09 Sep 2022 11:53
Operator : SY/VA
Sample : N4440-06
Misc : 5.05g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Sep 10 00:40:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Sep 10 00:38:15 2022
Response via : Initial Calibration

