

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091222\
 Data File : VW024602.D
 Acq On : 12 Sep 2022 12:37
 Operator : SY/VA
 Sample : VW0912SBL03
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0912SBL03

Quant Time: Sep 13 02:45:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Sep 13 02:43:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	214388	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	231398	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	93125	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.343	65	143283	32.371	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 129.480%	
7) Chloroethane-d5	2.873	69	113984	34.633	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 138.520%	
11) 1,1-Dichloroethene-d2	4.007	63	77829	16.666	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 66.680%	
21) 2-Butanone-d5	7.080	46	38474	53.882	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 107.760%	
24) Chloroform-d	7.641	84	132777	24.317	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 97.280%	
26) 1,2-Dichloroethane-d4	8.305	65	87811	28.523	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 114.080%	
32) Benzene-d6	8.275	84	263492	23.149	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 92.600%	
36) 1,2-Dichloropropane-d6	9.275	67	77679	23.147	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 92.600%	
41) Toluene-d8	10.323	98	254130	22.902	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 91.600%	
43) trans-1,3-Dichloroprop...	10.579	79	33258	22.114	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 88.440%	
47) 2-Hexanone-d5	10.921	63	34238	55.504	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 111.000%	
56) 1,1,2,2-Tetrachloroeth...	12.689	84	74687	23.220	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 92.880%	
66) 1,2-Dichlorobenzene-d4	13.853	152	78167	26.501	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 106.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091222\
Data File : VW024602.D
Acq On : 12 Sep 2022 12:37
Operator : SY/VA
Sample : VW0912SBL03
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0912SBL03

Quant Time: Sep 13 02:45:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM090622SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Sep 13 02:43:10 2022
Response via : Initial Calibration

