

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091725\
 Data File : VW032188.D
 Acq On : 17 Sep 2025 14:37
 Operator : SY/MD
 Sample : Q3126-01
 Misc : 5.09g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 OK-01-091725

Quant Time: Sep 18 04:35:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	142431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	307406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	310863	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	157856	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	125317	61.522	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	123.040%	
35) Dibromofluoromethane	7.904	113	107622	49.900	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	99.800%	
50) Toluene-d8	10.325	98	357784	47.256	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	94.520%	
62) 4-Bromofluorobenzene	12.617	95	142391	50.859	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	101.720%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.960	84	45570	22.762	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091725\
Data File : VW032188.D
Acq On : 17 Sep 2025 14:37
Operator : SY/MD
Sample : Q3126-01
Misc : 5.09g/5mL/MSVOA_W/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
OK-01-091725

Quant Time: Sep 18 04:35:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

