

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091825\
 Data File : VW032196.D
 Acq On : 18 Sep 2025 12:04
 Operator : SY/MD
 Sample : Q3118-01RE
 Misc : 5.08g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EO-01-091625RE

Quant Time: Sep 19 04:33:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.953	168	60924	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.849	114	126873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	114259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	49578	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	46374	53.224	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	106.440%
35) Dibromofluoromethane	7.898	113	41943	47.120	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	94.240%
50) Toluene-d8	10.331	98	128179	41.020	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	82.040%
62) 4-Bromofluorobenzene	12.617	95	43700	37.819	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	75.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091825\
Data File : VW032196.D
Acq On : 18 Sep 2025 12:04
Operator : SY/MD
Sample : Q3118-01RE
Misc : 5.08g/5mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EO-01-091625RE

Quant Time: Sep 19 04:33:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

