

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091825\
 Data File : VW032200.D
 Acq On : 18 Sep 2025 15:26
 Operator : SY/MD
 Sample : Q3133-07
 Misc : 3.45g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VNJ-210

Quant Time: Sep 19 04:35:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	131697	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	307869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	326005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.562	152	169360	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	127639	67.769	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	135.540%
35) Dibromofluoromethane	7.904	113	114892	53.191	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	106.380%
50) Toluene-d8	10.331	98	380512	50.182	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.617	95	148146	52.835	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	105.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091825\
Data File : VW032200.D
Acq On : 18 Sep 2025 15:26
Operator : SY/MD
Sample : Q3133-07
Misc : 3.45g/5mL/MSVOA_W/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VNJ-210

Quant Time: Sep 19 04:35:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

