



Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091825\
Data File : VW032191.D
Acq On : 18 Sep 2025 08:38
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Title : SW846 8260
Last Update : Wed Aug 27 04:13:40 2025



AutoFind: Scans 1759, 1760, 1761; Background Corrected with Scan 1750

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	21.8	16838	PASS
75	95	30	60	57.0	44067	PASS
95	95	100	100	100.0	77336	PASS
96	95	5	9	6.6	5077	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	75.1	58067	PASS
175	174	5	9	7.5	4348	PASS
176	174	95	101	95.9	55664	PASS
177	176	5	9	5.4	3025	PASS