

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
 Data File : VW032205.D
 Acq On : 19 Sep 2025 10:08
 Operator : SY/MD
 Sample : VW0919SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0919SBL01

Quant Time: Sep 19 23:04:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	128568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	281974	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	287563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	136277	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	107935	58.702	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	117.400%
35) Dibromofluoromethane	7.898	113	95858	48.454	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	96.900%
50) Toluene-d8	10.325	98	310434	44.700	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	89.400%
62) 4-Bromofluorobenzene	12.617	95	121634	47.364	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	94.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
Data File : VW032205.D
Acq On : 19 Sep 2025 10:08
Operator : SY/MD
Sample : VW0919SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0919SBL01

Quant Time: Sep 19 23:04:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

