

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
 Data File : VW032210.D
 Acq On : 19 Sep 2025 12:24
 Operator : SY/MD
 Sample : Q3140-01
 Misc : 5.14g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 TR-04-091825

Quant Time: Sep 19 23:08:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	138085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	308486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.636	117	326487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	167787	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	112937	57.189	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	=	114.380%	
35) Dibromofluoromethane	7.898	113	102964	47.573	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	=	95.140%	
50) Toluene-d8	10.325	98	360074	47.392	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	=	94.780%	
62) 4-Bromofluorobenzene	12.617	95	166720	59.341	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	=	118.680%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.954	84	29608	13.873	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
Data File : VW032210.D
Acq On : 19 Sep 2025 12:24
Operator : SY/MD
Sample : Q3140-01
Misc : 5.14g/5mL/MSVOA_W/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
TR-04-091825

Quant Time: Sep 19 23:08:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

