

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
 Data File : VW032212.D
 Acq On : 19 Sep 2025 13:59
 Operator : SY/MD
 Sample : Q3145-03
 Misc : 5.15g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 SP-B

Quant Time: Sep 19 23:09:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	157059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	343989	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	339827	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	168068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	120206	53.516	ug/l	0.00
Spiked Amount	50.000	Range 63 - 155	Recovery	= 107.040%		
35) Dibromofluoromethane	7.904	113	109313	45.294	ug/l	0.00
Spiked Amount	50.000	Range 70 - 134	Recovery	= 90.580%		
50) Toluene-d8	10.331	98	368459	43.490	ug/l	0.00
Spiked Amount	50.000	Range 74 - 123	Recovery	= 86.980%		
62) 4-Bromofluorobenzene	12.617	95	154079	49.181	ug/l	0.00
Spiked Amount	50.000	Range 17 - 146	Recovery	= 98.360%		
Target Compounds						
20) Methylene Chloride	4.948	84	20076	6.674	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW091925\
Data File : VW032212.D
Acq On : 19 Sep 2025 13:59
Operator : SY/MD
Sample : Q3145-03
Misc : 5.15g/5mL/MSVOA_W/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
SP-B

Quant Time: Sep 19 23:09:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

