

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092324\
 Data File : VW030206.D
 Acq On : 23 Sep 2024 17:41
 Operator : SY/MD
 Sample : P4131-08
 Misc : 5.25g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EW9

Quant Time: Sep 24 01:20:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Sep 24 01:15:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	488203	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	300632	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	75751	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	135723	21.527	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.120%
7) Chloroethane-d5	2.899	69	113827	25.573	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.280%
11) 1,1-Dichloroethene-d2	4.021	65	54792	17.393	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.560%
21) 2-Butanone-d5	7.075	46	32451	21.257	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	42.520%
24) Chloroform-d	7.648	84	257640	19.193	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.760%
26) 1,2-Dichloroethane-d4	8.307	65	143539	21.294	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.160%
32) Benzene-d6	8.276	84	486567	28.094	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	112.360%
36) 1,2-Dichloropropane-d6	9.276	67	150275	28.103	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	112.400%
41) Toluene-d8	10.318	98	354441	22.837	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.360%
43) trans-1,3-Dichloroprop...	10.574	79	28989	13.532	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	54.120%
47) 2-Hexanone-d5	10.922	63	25325	30.950	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.900%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	81358	20.436	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	81.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	47591	18.507	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	74.040%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092324\
Data File : VW030206.D
Acq On : 23 Sep 2024 17:41
Operator : SY/MD
Sample : P4131-08
Misc : 5.25g/10mL/MSVOA_W/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EW9

Quant Time: Sep 24 01:20:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Sep 24 01:15:59 2024
Response via : Initial Calibration

