

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
 Data File : VW030217.D
 Acq On : 24 Sep 2024 12:04
 Operator : SY/MD
 Sample : VIBLK526
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK526

Quant Time: Sep 25 01:19:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 25 01:17:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	307148	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	287167	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	132950	25.000	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.362	65	85831	21.639	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.560%
7) Chloroethane-d5	2.899	69	72780	25.990	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	103.960%
11) 1,1-Dichloroethene-d2	4.021	65	34666	17.491	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	69.960%
21) 2-Butanone-d5	7.087	46	36738	38.251	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	76.500%
24) Chloroform-d	7.648	84	171528	20.311	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.240%
26) 1,2-Dichloroethane-d4	8.307	65	92047	21.705	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.800%
32) Benzene-d6	8.270	84	329037	19.889	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.560%
36) 1,2-Dichloropropane-d6	9.276	67	98411	19.267	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.080%
41) Toluene-d8	10.318	98	297372	20.059	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.240%
43) trans-1,3-Dichloroprop...	10.574	79	39648	19.375	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	77.480%
47) 2-Hexanone-d5	10.922	63	30499	39.021	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.040%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	75935	19.968	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	79.880%
66) 1,2-Dichlorobenzene-d4	13.848	152	99467	22.039	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
Data File : VW030217.D
Acq On : 24 Sep 2024 12:04
Operator : SY/MD
Sample : VIBLK526
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK526

Quant Time: Sep 25 01:19:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 25 01:17:29 2024
Response via : Initial Calibration

