

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
 Data File : VW030225.D
 Acq On : 24 Sep 2024 15:39
 Operator : SY/MD
 Sample : VIBLK527
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK527

Quant Time: Sep 25 01:21:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 25 01:17:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	256167	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	239719	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	107378	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	68660	20.755	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.000%
7) Chloroethane-d5	2.899	69	59396	25.432	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.720%
11) 1,1-Dichloroethene-d2	4.027	65	29133	17.625	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	70.520%
21) 2-Butanone-d5	7.081	46	34304	42.824	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	85.640%
24) Chloroform-d	7.648	84	140832	19.995	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.960%
26) 1,2-Dichloroethane-d4	8.307	65	76652	21.672	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.680%
32) Benzene-d6	8.276	84	271003	19.624	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.480%
36) 1,2-Dichloropropane-d6	9.276	67	82910	19.445	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	77.760%
41) Toluene-d8	10.325	98	247131	19.969	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	79.880%
43) trans-1,3-Dichloroprop...	10.575	79	31686	18.549	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.200%
47) 2-Hexanone-d5	10.922	63	27535	42.202	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.400%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	66490	20.946	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	13.849	152	83590	22.932	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
Data File : VW030225.D
Acq On : 24 Sep 2024 15:39
Operator : SY/MD
Sample : VIBLK527
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK527

Quant Time: Sep 25 01:21:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 25 01:17:29 2024
Response via : Initial Calibration

