

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
 Data File : VW030227.D
 Acq On : 24 Sep 2024 16:26
 Operator : SY/MD
 Sample : VIBLK528
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK528

Quant Time: Sep 25 01:22:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 25 01:17:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	257589	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	252200	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	117259	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	76853	23.103	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.400%
7) Chloroethane-d5	2.899	69	65825	28.029	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.120%
11) 1,1-Dichloroethene-d2	4.015	65	31207	18.776	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.120%
21) 2-Butanone-d5	7.081	46	41212	51.164	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.320%
24) Chloroform-d	7.648	84	153080	21.614	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.440%
26) 1,2-Dichloroethane-d4	8.307	65	84443	23.742	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.960%
32) Benzene-d6	8.276	84	296254	20.391	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	81.560%
36) 1,2-Dichloropropane-d6	9.270	67	91073	20.302	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.200%
41) Toluene-d8	10.325	98	267018	20.508	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	82.040%
43) trans-1,3-Dichloroprop...	10.575	79	36033	20.050	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.200%
47) 2-Hexanone-d5	10.922	63	34134	49.727	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	99.460%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	78521	23.511	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	94.040%
66) 1,2-Dichlorobenzene-d4	13.849	152	94223	23.671	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092424\
Data File : VW030227.D
Acq On : 24 Sep 2024 16:26
Operator : SY/MD
Sample : VIBLK528
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK528

Quant Time: Sep 25 01:22:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 25 01:17:29 2024
Response via : Initial Calibration

