

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
 Data File : VW032253.D
 Acq On : 24 Sep 2025 18:41
 Operator : SY/MD
 Sample : Q3183-09
 Misc : 3.82g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 WC-3-VOC

Quant Time: Sep 25 08:17:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	148144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	302431	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	303302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	158964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	138551	65.396	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	130.800%
35) Dibromofluoromethane	7.904	113	112473	53.007	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	106.020%
50) Toluene-d8	10.331	98	382494	51.351	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	102.700%
62) 4-Bromofluorobenzene	12.617	95	132962	48.273	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	96.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
Data File : VW032253.D
Acq On : 24 Sep 2025 18:41
Operator : SY/MD
Sample : Q3183-09
Misc : 3.82g/5mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
WC-3-VOC

Quant Time: Sep 25 08:17:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

