

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
 Data File : VW032256.D
 Acq On : 24 Sep 2025 19:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 08:19:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.330	0.261	20.9	75	0.00
3 P	Chloromethane	0.346	0.302	12.7	79	0.00
4 C	Vinyl Chloride	0.450	0.409	9.1#	82	0.00
5 T	Bromomethane	0.401	0.355	11.5	84	0.00
6 T	Chloroethane	0.317	0.304	4.1	85	0.00
7 T	Trichlorofluoromethane	0.445	0.415	6.7	77	0.00
8 T	Diethyl Ether	0.332	0.331	0.3	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.489	2.6	86	0.00
10 T	Methyl Iodide	0.764	0.762	0.3	91	0.00
11 T	Tert butyl alcohol	0.044	0.042	4.5	84	0.00
12 CM	1,1-Dichloroethene	0.528	0.524	0.8#	87	0.00
13 T	Acrolein	0.067	0.007	89.6#	10#	0.00
14 T	Allyl chloride	0.751	0.725	3.5	84	0.00
15 T	Acrylonitrile	0.162	0.177	-9.3	96	0.00
16 T	Acetone	0.155	0.134	13.5	76	0.00
17 T	Carbon Disulfide	1.322	1.159	12.3	76	0.00
18 T	Methyl Acetate	0.493	0.727	-47.5#	126	0.00
19 T	Methyl tert-butyl Ether	0.998	1.093	-9.5	93	0.00
20 T	Methylene Chloride	0.725	0.668	7.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.599	-1.5	88	0.00
22 T	Diisopropyl ether	1.634	1.769	-8.3	92	0.00
23 T	Vinyl Acetate	1.111	1.061	4.5	80	0.00
24 P	1,1-Dichloroethane	1.070	1.126	-5.2	91	0.00
25 T	2-Butanone	0.219	0.235	-7.3	92	0.00
26 T	2,2-Dichloropropane	0.566	0.526	7.1	77	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.763	-6.3	92	0.00
28 T	Bromochloromethane	0.490	0.513	-4.7	94	0.00
29 T	Tetrahydrofuran	0.139	0.153	-10.1	96	0.00
30 C	Chloroform	1.221	1.310	-7.3#	93	0.00
31 T	Cyclohexane	0.878	0.781	11.0	81	0.00
32 T	1,1,1-Trichloroethane	0.891	0.921	-3.4	89	0.00
33 S	1,2-Dichloroethane-d4	0.715	0.743	-3.9	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.351	0.362	-3.1	93	0.00
36 T	1,1-Dichloropropene	0.455	0.474	-4.2	86	0.00
37 T	Ethyl Acetate	0.273	0.278	-1.8	87	0.00
38 T	Carbon Tetrachloride	0.488	0.496	-1.6	84	0.00
39 T	Methylcyclohexane	0.539	0.541	-0.4	81	0.00
40 TM	Benzene	1.406	1.521	-8.2	92	0.00
41 T	Methacrylonitrile	0.152	0.173	-13.8	102	0.00
42 TM	1,2-Dichloroethane	0.487	0.536	-10.1	94	0.00
43 T	Isopropyl Acetate	0.512	0.538	-5.1	88	0.00
44 TM	Trichloroethene	0.354	0.381	-7.6	94	0.00
45 C	1,2-Dichloropropane	0.339	0.372	-9.7#	93	0.00
46 T	Dibromomethane	0.238	0.265	-11.3	98	0.00
47 T	Bromodichloromethane	0.540	0.592	-9.6	92	0.00
48 T	Methyl methacrylate	0.236	0.260	-10.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
 Data File : VW032256.D
 Acq On : 24 Sep 2025 19:46
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_W
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 25 08:19:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 27 04:13:40 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	91	0.00
50 S	Toluene-d8	1.231	1.275	-3.6	89	0.00
51 T	4-Methyl-2-Pentanone	0.289	0.331	-14.5	97	0.00
52 CM	Toluene	0.919	0.980	-6.6#	90	0.00
53 T	t-1,3-Dichloropropene	0.477	0.510	-6.9	88	0.00
54 T	cis-1,3-Dichloropropene	0.544	0.579	-6.4	87	0.00
55 T	1,1,2-Trichloroethane	0.309	0.350	-13.3	101	0.00
56 T	Ethyl methacrylate	0.395	0.455	-15.2	93	0.00
57 T	1,3-Dichloropropane	0.518	0.576	-11.2	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.211	0.251	-19.0	104	0.00
59 T	2-Hexanone	0.201	0.231	-14.9	96	0.00
60 T	Dibromochloromethane	0.374	0.419	-12.0	93	0.00
61 T	1,2-Dibromoethane	0.304	0.342	-12.5	95	0.00
62 S	4-Bromofluorobenzene	0.455	0.479	-5.3	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.329	0.365	-10.9	97	0.00
65 PM	Chlorobenzene	1.138	1.184	-4.0	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.375	0.415	-10.7	97	0.00
67 C	Ethyl Benzene	1.865	2.009	-7.7#	93	0.00
68 T	m/p-Xylenes	0.718	0.775	-7.9	91	0.00
69 T	o-Xylene	0.679	0.733	-8.0	92	0.00
70 T	Styrene	1.197	1.339	-11.9	95	0.00
71 P	Bromoform	0.223	0.246	-10.3	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	3.397	3.453	-1.6	90	0.00
74 T	N-amyl acetate	0.986	0.945	4.2	87	0.00
75 P	1,1,2,2-Tetrachloroethane	0.807	0.832	-3.1	97	0.00
76 T	1,2,3-Trichloropropane	0.626	0.655	-4.6	102	0.00
77 T	Bromobenzene	0.866	0.916	-5.8	96	0.00
78 T	n-propylbenzene	4.124	4.258	-3.2	93	0.00
79 T	2-Chlorotoluene	2.522	2.592	-2.8	93	0.00
80 T	1,3,5-Trimethylbenzene	2.876	2.969	-3.2	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.233	0.229	1.7	87	0.00
82 T	4-Chlorotoluene	2.699	2.779	-3.0	92	0.00
83 T	tert-Butylbenzene	2.439	2.601	-6.6	91	0.00
84 T	1,2,4-Trimethylbenzene	2.975	3.019	-1.5	89	0.00
85 T	sec-Butylbenzene	3.641	3.730	-2.4	89	0.00
86 T	p-Isopropyltoluene	3.096	3.175	-2.6	89	0.00
87 T	1,3-Dichlorobenzene	1.744	1.811	-3.8	94	0.00
88 T	1,4-Dichlorobenzene	1.748	1.818	-4.0	98	0.00
89 T	n-Butylbenzene	2.952	3.044	-3.1	90	0.00
90 T	Hexachloroethane	0.566	0.570	-0.7	93	0.00
91 T	1,2-Dichlorobenzene	1.544	1.654	-7.1	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.143	0.144	-0.7	97	0.00
93 T	1,2,4-Trichlorobenzene	0.987	0.974	1.3	89	0.00
94 T	Hexachlorobutadiene	0.489	0.463	5.3	83	0.00
95 T	Naphthalene	2.233	2.473	-10.7	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092425\
Data File : VW032256.D
Acq On : 24 Sep 2025 19:46
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_W
LabSampleId :
VSTDCCC050

Quant Time: Sep 25 08:19:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W082625S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 27 04:13:40 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.918	0.920	-0.2	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6