

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092524\
 Data File : VW030238.D
 Acq On : 25 Sep 2024 11:28
 Operator : SY/MD
 Sample : VW0925SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK557

Quant Time: Sep 26 01:19:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Sep 26 01:18:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	198575	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	188822	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	83704	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	56864	22.174	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.680%
7) Chloroethane-d5	2.905	69	45325	25.035	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	100.160%
11) 1,1-Dichloroethene-d2	4.027	65	26569	20.736	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	82.960%
21) 2-Butanone-d5	7.081	46	31666	50.996	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.000%
24) Chloroform-d	7.648	84	134353	24.607	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.440%
26) 1,2-Dichloroethane-d4	8.307	65	72045	26.277	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	105.120%
32) Benzene-d6	8.276	84	260157	23.916	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.680%
36) 1,2-Dichloropropane-d6	9.270	67	81978	24.409	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	97.640%
41) Toluene-d8	10.325	98	225318	23.114	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.440%
43) trans-1,3-Dichloroprop...	10.581	79	29388	21.841	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.360%
47) 2-Hexanone-d5	10.922	63	23898	46.501	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	59653	23.857	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.440%
66) 1,2-Dichlorobenzene-d4	13.849	152	72048	25.356	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092524\
Data File : VW030238.D
Acq On : 25 Sep 2024 11:28
Operator : SY/MD
Sample : VW0925SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK557

Quant Time: Sep 26 01:19:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM091624SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Sep 26 01:18:41 2024
Response via : Initial Calibration

