

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
 Data File : VW030255.D
 Acq On : 26 Sep 2024 12:42
 Operator : SY/MD
 Sample : VIBLK529
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK529

Quant Time: Sep 27 04:31:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 27 04:25:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	175624	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	164719	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	75168	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	59173	24.971	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	99.880%		
7) Chloroethane-d5	2.899	69	39422	22.570	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	90.280%		
11) 1,1-Dichloroethene-d2	4.021	65	29215	24.888	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	99.560%		
21) 2-Butanone-d5	7.075	46	29050	51.196	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	102.400%		
24) Chloroform-d	7.648	84	121351	23.336	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	93.360%		
26) 1,2-Dichloroethane-d4	8.307	65	67657	25.508	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	102.040%		
32) Benzene-d6	8.276	84	250473	24.072	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	96.280%		
36) 1,2-Dichloropropane-d6	9.276	67	77021	23.909	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	95.640%		
41) Toluene-d8	10.324	98	225211	24.354	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	97.400%		
43) trans-1,3-Dichloroprop...	10.581	79	30025	24.590	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	98.360%		
47) 2-Hexanone-d5	10.922	63	22015	48.821	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	97.640%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	53297	24.221	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	96.880%		
66) 1,2-Dichlorobenzene-d4	13.848	152	69032	26.390	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	105.560%		
Target Compounds						Qvalue
70) 1,2,4-trichlorobenzene	15.128	180	3517	1.361	ug/L	97
72) 1,2,3-Trichlorobenzene	15.555	180	3892	1.728	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
Data File : VW030255.D
Acq On : 26 Sep 2024 12:42
Operator : SY/MD
Sample : VIBLK529
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK529

Quant Time: Sep 27 04:31:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 27 04:25:21 2024
Response via : Initial Calibration

