

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
 Data File : VW030263.D
 Acq On : 26 Sep 2024 16:55
 Operator : SY/MD
 Sample : P4183-05
 Misc : 4.80g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EM1

Quant Time: Sep 27 04:40:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 27 04:39:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	277790	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	177291	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	48022	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	41018	10.943	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	43.760%
7) Chloroethane-d5	2.899	69	35037	12.682	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.720%
11) 1,1-Dichloroethene-d2	4.021	65	21814	11.749	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	47.000%
21) 2-Butanone-d5	7.075	46	18746	20.887	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	41.780%
24) Chloroform-d	7.648	84	116477	14.161	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	56.640%
26) 1,2-Dichloroethane-d4	8.307	65	68176	16.250	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	65.000%#
32) Benzene-d6	8.276	84	207485	18.527	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.120%
36) 1,2-Dichloropropane-d6	9.276	67	70150	20.232	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	80.920%
41) Toluene-d8	10.319	98	151857	15.257	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	61.040%
43) trans-1,3-Dichloroprop...	10.575	79	19328	14.707	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	58.840%
47) 2-Hexanone-d5	10.922	63	16051	33.071	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	66.140%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	40855	17.250	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	69.000%
66) 1,2-Dichlorobenzene-d4	13.848	152	22725	13.598	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	54.400%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
Data File : VW030263.D
Acq On : 26 Sep 2024 16:55
Operator : SY/MD
Sample : P4183-05
Misc : 4.80g/10mL/MSVOA_W/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EM1

Quant Time: Sep 27 04:40:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 27 04:39:02 2024
Response via : Initial Calibration

