

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
 Data File : VW030276.D
 Acq On : 26 Sep 2024 22:07
 Operator : SY/MD
 Sample : P4183-21
 Misc : 5.65g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EM9

Quant Time: Sep 27 04:44:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 27 04:39:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	641631	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	549044	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	238232	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	60852	7.029	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	28.120%#
7) Chloroethane-d5	2.899	69	50125	7.855	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	31.400%
11) 1,1-Dichloroethene-d2	4.021	65	40026	9.333	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	37.320%#
21) 2-Butanone-d5	7.081	46	51787	24.981	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	49.960%
24) Chloroform-d	7.648	84	213703	11.248	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	45.000%
26) 1,2-Dichloroethane-d4	8.307	65	110552	11.408	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	45.640%#
32) Benzene-d6	8.276	84	425698	12.274	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	49.080%
36) 1,2-Dichloropropane-d6	9.276	67	134779	12.552	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	50.200%#
41) Toluene-d8	10.319	98	375094	12.169	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	48.680%
43) trans-1,3-Dichloroprop...	10.575	79	49535	12.171	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	48.680%
47) 2-Hexanone-d5	10.922	63	46348	30.836	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.680%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	98902	13.484	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	53.920%
66) 1,2-Dichlorobenzene-d4	13.848	152	114059	13.758	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	55.040%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092624\
Data File : VW030276.D
Acq On : 26 Sep 2024 22:07
Operator : SY/MD
Sample : P4183-21
Misc : 5.65g/10mL/MSVOA_W/SOIL/A
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EM9

Quant Time: Sep 27 04:44:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092624SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 27 04:39:02 2024
Response via : Initial Calibration

