

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092625\
 Data File : VW032269.D
 Acq On : 26 Sep 2025 10:08
 Operator : SY/MD
 Sample : VW0926SBS01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01

Quant Time: Sep 27 04:31:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh
 Dadoda
 09/29/2025
 Supervised By :Semsettin
 Yesilyurt
 09/29/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.959	168	258288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	477257	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	426681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	223480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	192065	47.094	ug/l	0.00
Spiked Amount 50.000	Range 63 - 155		Recovery =	94.180%		
35) Dibromofluoromethane	7.904	113	158775	46.259	ug/l	0.00
Spiked Amount 50.000	Range 70 - 134		Recovery =	92.520%		
50) Toluene-d8	10.325	98	579559	46.852	ug/l	0.00
Spiked Amount 50.000	Range 74 - 123		Recovery =	93.700%		
62) 4-Bromofluorobenzene	12.617	95	225595	48.349	ug/l	0.00
Spiked Amount 50.000	Range 17 - 146		Recovery =	96.700%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.046	85	37842	21.206	ug/l	98
3) Chloromethane	2.253	50	51625	20.215	ug/l	96
4) Vinyl Chloride	2.405	62	67947	22.772	ug/l	96
5) Bromomethane	2.820	94	47148	23.140	ug/l	88
6) Chloroethane	2.966	64	43409	21.469	ug/l	96
7) Trichlorofluoromethane	3.301	101	41970	18.668	ug/l	98
8) Diethyl Ether	3.716	74	42168	19.836	ug/l	100
9) 1,1,2-Trichlorotrifluo...	4.100	101	62237	21.687	ug/l	100
10) Methyl Iodide	4.307	142	83781	21.585	ug/l	99
11) Tert butyl alcohol	5.222	59	30277	106.950	ug/l	96
12) 1,1-Dichloroethene	4.082	96	67892	22.132	ug/l	87
13) Acrolein	3.929	56	48232	95.724	ug/l	96
14) Allyl chloride	4.698	41	101529	21.326	ug/l	99
15) Acrylonitrile	5.399	53	99331	90.724	ug/l	98
16) Acetone	4.161	43	124429	118.802	ug/l	100
17) Carbon Disulfide	4.417	76	156951	22.253	ug/l	99
18) Methyl Acetate	4.704	43	58117	16.134	ug/l	97
19) Methyl tert-butyl Ether	5.466	73	116547	20.496	ug/l	98
20) Methylene Chloride	4.954	84	97152	21.347	ug/l	97
21) trans-1,2-Dichloroethene	5.460	96	69980	21.155	ug/l	96
22) Diisopropyl ether	6.338	45	224780	21.248	ug/l	93
23) Vinyl Acetate	6.283	43	702126	103.170	ug/l	99
24) 1,1-Dichloroethane	6.240	63	139846	20.941	ug/l	99
25) 2-Butanone	7.197	43	147498	101.035	ug/l	93
26) 2,2-Dichloropropane	7.185	77	69568	21.140	ug/l	100
27) cis-1,2-Dichloroethene	7.191	96	83416	20.275	ug/l	97
28) Bromochloromethane	7.532	49	64260	20.166	ug/l #	99
29) Tetrahydrofuran	7.551	42	85741	90.468	ug/l	99
30) Chloroform	7.691	83	144668	20.763	ug/l	98
31) Cyclohexane	7.971	56	112115	21.291	ug/l	95
32) 1,1,1-Trichloroethane	7.886	97	99064	21.104	ug/l	98
36) 1,1-Dichloropropene	8.093	75	94168	20.350	ug/l	98
37) Ethyl Acetate	7.270	43	59573	19.411	ug/l	99
38) Carbon Tetrachloride	8.081	117	90875	20.874	ug/l	99
39) Methylcyclohexane	9.343	83	108509	20.265	ug/l	97
40) Benzene	8.337	78	305208	20.954	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092625\
 Data File : VW032269.D
 Acq On : 26 Sep 2025 10:08
 Operator : SY/MD
 Sample : VW0926SBS01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01

Quant Time: Sep 27 04:31:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh
 Dadoda
 09/29/2025
 Supervised By :Semsettin
 Yesilyurt
 09/29/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.496	41	31816	18.101	ug/l	96
42) 1,2-Dichloroethane	8.410	62	97519	20.142	ug/l	99
43) Isopropyl Acetate	8.435	43	103104	18.768	ug/l	99
44) Trichloroethene	9.099	130	71206	20.915	ug/l	94
45) 1,2-Dichloropropane	9.374	63	77145	20.449	ug/l	98
46) Dibromomethane	9.465	93	46372	19.993	ug/l	98
47) Bromodichloromethane	9.648	83	106292	20.093	ug/l	100
48) Methyl methacrylate	9.441	41	48310	18.692	ug/l	98
49) 1,4-Dioxane	9.459	88	11583	372.877	ug/l #	94
51) 4-Methyl-2-Pentanone	10.215	43	299407	91.920	ug/l	98
52) Toluene	10.392	92	187424	20.760	ug/l	99
53) t-1,3-Dichloropropene	10.611	75	97505	20.117	ug/l	95
54) cis-1,3-Dichloropropene	10.075	75	114169	20.403	ug/l	98
55) 1,1,2-Trichloroethane	10.788	97	61241	19.344	ug/l	94
56) Ethyl methacrylate	10.648	69	79427	19.046	ug/l	99
57) 1,3-Dichloropropane	10.934	76	108013	19.749	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.928	63	200248	87.473	ug/l	100
59) 2-Hexanone	10.971	43	213695	96.071	ug/l	97
60) Dibromochloromethane	11.129	129	69693	20.044	ug/l	98
61) 1,2-Dibromoethane	11.239	107	59022	19.678	ug/l	97
64) Tetrachloroethene	10.861	164	55720	21.429	ug/l	86
65) Chlorobenzene	11.660	112	213331	21.706	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.727	131	64799	20.847	ug/l	97
67) Ethyl Benzene	11.727	91	349768	21.742	ug/l	100
68) m/p-Xylenes	11.837	106	272503	44.189	ug/l	97
69) o-Xylene	12.166	106	124923	21.555	ug/l	96
70) Styrene	12.178	104	225931	22.230	ug/l	100
71) Bromoform	12.349	173	35746	19.735	ug/l	97
73) Isopropylbenzene	12.464	105	320847	20.627	ug/l	100
74) N-amyl acetate	12.269	43	93561	19.100	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.714	83	76820	18.837	ug/l	99
76) 1,2,3-Trichloropropane	12.763	75	56341m	18.325	ug/l	
77) Bromobenzene	12.745	156	77712	20.312	ug/l	93
78) n-propylbenzene	12.800	91	415955	21.290	ug/l	100
79) 2-Chlorotoluene	12.891	91	250287	21.147	ug/l	100
80) 1,3,5-Trimethylbenzene	12.940	105	275113	20.687	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.513	75	21754	18.378	ug/l	93
82) 4-Chlorotoluene	12.989	91	260792	20.849	ug/l	100
83) tert-Butylbenzene	13.202	119	224202	20.159	ug/l	98
84) 1,2,4-Trimethylbenzene	13.245	105	278390	20.618	ug/l	99
85) sec-Butylbenzene	13.379	105	347938	20.833	ug/l	99
86) p-Isopropyltoluene	13.495	119	290111	20.974	ug/l	99
87) 1,3-Dichlorobenzene	13.495	146	163503	21.606	ug/l	98
88) 1,4-Dichlorobenzene	13.574	146	159112	20.581	ug/l	98
89) n-Butylbenzene	13.818	91	289490	20.855	ug/l	99
90) Hexachloroethane	14.092	117	53545	20.770	ug/l	97
91) 1,2-Dichlorobenzene	13.867	146	147974	20.903	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	14.476	75	12549	17.990	ug/l	94
93) 1,2,4-Trichlorobenzene	15.123	180	83072	19.780	ug/l	100
94) Hexachlorobutadiene	15.232	225	41545	21.539	ug/l	87
95) Naphthalene	15.360	128	188745	18.725	ug/l	99
96) 1,2,3-Trichlorobenzene	15.543	180	74748	19.141	ug/l	93

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092625\
Data File : VW032269.D
Acq On : 26 Sep 2025 10:08
Operator : SY/MD
Sample : VW0926SBS01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Sep 27 04:31:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025

Supervised By :Semsettin
Yesilyurt
09/29/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092625\
Data File : VW032269.D
Acq On : 26 Sep 2025 10:08
Operator : SY/MD
Sample : VW0926SBS01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Quant Time: Sep 27 04:31:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh


Dadoda


09/29/2025

Supervised By :Semsetin

Yesilyurt


09/29/2025





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.959 min Scan# 996
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:168 Resp: 25828
Ion Ratio Lower Upper
168 100
99 67.7 55.0 82.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#2
Dichlorodifluoromethane
Concen: 21.206 ug/l
RT: 2.046 min Scan# 26
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 85 Resp: 37842
Ion Ratio Lower Upper
85 100
87 34.8 16.7 50.1





#3
Chloromethane
Concen: 20.215 ug/l
RT: 2.253 min Scan# 60
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01





Tgt Ion: 50 Resp: 51629
Ion Ratio Lower Upper
50 100
52 33.7 28.9 43.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#4
Vinyl Chloride
Concen: 22.772 ug/l
RT: 2.405 min Scan# 85
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 62 Resp: 67947
Ion Ratio Lower Upper
62 100
64 30.7 26.2 39.4





#5
Bromomethane
Concen: 23.140 ug/l
RT: 2.820 min Scan# 153
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01



Tgt Ion: 94 Resp: 4714
Ion Ratio Lower Upper
94 100
96 98.7 70.0 105.0




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025



#6
Chloroethane
Concen: 21.469 ug/l
RT: 2.966 min Scan# 177
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 64 Resp: 43409
Ion Ratio Lower Upper
64 100
66 29.4 25.5 38.3





#7
 Trichlorofluoromethane
 Concen: 18.668 ug/l
 RT: 3.301 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08





Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01

Tgt Ion:101 Resp: 41970
 Ion Ratio Lower Upper
 101 100
 103 64.1 50.3 75.5

Manual Integrations
 APPROVED




Reviewed By :Mahesh
 Dadoda
 09/29/2025


Supervised By :Semsettin
 Yesilyurt
 09/29/2025



#8
 Diethyl Ether
 Concen: 19.836 ug/l
 RT: 3.716 min Scan# 300
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Tgt Ion: 74 Resp: 42168
 Ion Ratio Lower Upper
 74 100
 45 98.0 49.0 147.2





#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.687 ug/l
RT: 4.100 min Scan# 363
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08





Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion:101 Resp: 6223
Ion Ratio Lower Upper
101 100
85 42.7 33.7 50.5
151 69.5 55.7 83.5

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#10
Methyl Iodide
Concen: 21.585 ug/l
RT: 4.307 min Scan# 397
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:142 Resp: 83781
Ion Ratio Lower Upper
142 100
127 46.6 36.9 55.3
141 14.9 12.0 18.0





#11
Tert butyl alcohol
Concen: 106.950 ug/l
RT: 5.222 min Scan# 546
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 59 Resp: 3027
Ion Ratio Lower Upper
59 100
57 9.5 8.9 13.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#12
1,1-Dichloroethene
Concen: 22.132 ug/l
RT: 4.082 min Scan# 360
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 96 Resp: 67892
Ion Ratio Lower Upper
96 100
61 148.8 136.3 204.5
98 60.3 50.7 76.1





#13
Acrolein
Concen: 95.724 ug/l
RT: 3.929 min Scan# 335
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01





Tgt Ion: 56 Resp: 4823
Ion Ratio Lower Upper
56 100
55 69.8 53.2 79.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#14
Allyl chloride
Concen: 21.326 ug/l
RT: 4.698 min Scan# 461
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 41 Resp: 101529
Ion Ratio Lower Upper
41 100
39 68.7 54.3 81.5
76 35.7 28.7 43.1





#15
Acrylonitrile
Concen: 90.724 ug/l
RT: 5.399 min Scan# 51
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 53 Resp: 9933
Ion Ratio Lower Upper
53 100
52 84.9 66.3 99.5
51 38.2 29.6 44.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#16
Acetone
Concen: 118.802 ug/l
RT: 4.161 min Scan# 373
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 43 Resp: 124429
Ion Ratio Lower Upper
43 100
58 31.5 25.4 38.0





#17

Carbon Disulfide

Concen: 22.253 ug/l

RT: 4.417 min Scan# 415

Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Instrument :

MSVOA_W

ClientSampleId :

VW0926SBS01

Tgt Ion: 76 Resp: 15695

Ion Ratio Lower Upper

76 100

78 8.6 7.2 10.8

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

09/29/2025

Supervised By :Semsettin

Yesilyurt

09/29/2025



#18

Methyl Acetate

Concen: 16.134 ug/l

RT: 4.704 min Scan# 462

Delta R.T. -0.006 min

Lab File: VW032269.D




Acq: 26 Sep 2025 10:08


Tgt Ion: 43 Resp: 58117

Ion Ratio Lower Upper

43 100


74 24.8 18.8 28.2





#19
Methyl tert-butyl Ether
Concen: 20.496 ug/l
RT: 5.466 min Scan# 586
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 73 Resp: 11654
Ion Ratio Lower Upper
73 100
57 22.6 17.3 25.9

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#20
Methylene Chloride
Concen: 21.347 ug/l
RT: 4.954 min Scan# 503
Delta R.T. -0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 84 Resp: 97152
Ion Ratio Lower Upper
84 100
49 126.4 97.8 146.6
51 38.7 29.9 44.9
86 64.5 51.3 76.9





#21
trans-1,2-Dichloroethene
Concen: 21.155 ug/l
RT: 5.460 min Scan# 585
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 96 Resp: 69980
Ion Ratio Lower Upper
96 100
61 142.6 119.2 178.8
98 63.1 52.5 78.7

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#22
Diisopropyl ether
Concen: 21.248 ug/l
RT: 6.338 min Scan# 730
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 45 Resp: 224780
Ion Ratio Lower Upper
45 100
43 64.7 46.2 69.4
87 25.3 22.2 33.2
59 13.2 9.5 14.3





#23
 Vinyl Acetate
 Concen: 103.170 ug/l
 RT: 6.283 min Scan# 713
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01





Tgt Ion: 43 Resp: 702120
 Ion Ratio Lower Upper
 43 100
 86 10.9 9.0 13.6


Manual Integrations
 APPROVED

Reviewed By :Mahesh
 Dadoda
 09/29/2025



Supervised By :Semsettin
 Yesilyurt
 09/29/2025




#24
 1,1-Dichloroethane
 Concen: 20.941 ug/l
 RT: 6.240 min Scan# 714
 Delta R.T. -0.006 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08



Tgt Ion: 63 Resp: 139846
 Ion Ratio Lower Upper
 63 100
 98 6.6 3.6 10.9
 100 4.8 2.4 7.1





#25
2-Butanone
Concen: 101.035 ug/l
RT: 7.197 min Scan# 871
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08





Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion: 43 Resp: 147498
Ion Ratio Lower Upper
43 100
72 24.3 22.3 33.5

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#26
2,2-Dichloropropane
Concen: 21.140 ug/l
RT: 7.185 min Scan# 869
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 77 Resp: 69568
Ion Ratio Lower Upper
77 100
97 23.4 11.7 35.0





#27

cis-1,2-Dichloroethene

Concen: 20.275 ug/l

RT: 7.191 min Scan# 870

Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Instrument :

MSVOA_W

ClientSampleId :

VW0926SBS01

Tgt Ion: 96	Resp: 83410
Ion Ratio	Lower Upper
96 100	
61 146.0	0.0 281.8
98 65.0	0.0 130.8

Manual Integrations
APPROVED

Reviewed By :Mahesh





Dadoda

09/29/2025

Supervised By :Semsettin

Yesilyurt

09/29/2025



#28

Bromochloromethane

Concen: 20.166 ug/l




RT: 7.532 min Scan# 926


Delta R.T. 0.006 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Tgt Ion: 49	Resp: 64260
Ion Ratio	Lower Upper
49 100	
129 1.5	0.0 2.2
130 71.0	0.0 0.0#





#29

Tetrahydrofuran

Concen: 90.468 ug/l

RT: 7.551 min Scan# 91

Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Instrument :

MSVOA_W

ClientSampleId :

VW0926SBS01

Tgt Ion: 42	Resp: 8574
Ion Ratio	Lower Upper
42 100	
72 44.1	35.1 52.7
71 40.5	33.3 49.9

Manual Integrations
APPROVED

Reviewed By :Mahesh





Dadoda

09/29/2025

Supervised By :Semsettin

Yesilyurt

09/29/2025



#30

Chloroform

Concen: 20.763 ug/l




RT: 7.691 min Scan# 952


Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08


Tgt Ion: 83	Resp: 144668
Ion Ratio	Lower Upper
83 100	
85 65.3	53.3 79.9





#31
Cyclohexane
Concen: 21.291 ug/l
RT: 7.971 min Scan# 998
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 56 Resp: 112115
Ion Ratio Lower Upper
56 100
69 35.1 28.2 42.2
84 83.2 71.8 107.6

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#32
1,1,1-Trichloroethane
Concen: 21.104 ug/l
RT: 7.886 min Scan# 984
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 97 Resp: 99064
Ion Ratio Lower Upper
97 100
99 61.4 50.6 76.0
61 45.7 36.2 54.4





#33
1,2-Dichloroethane-d4
Concen: 47.094 ug/l
RT: 8.313 min Scan# 1054
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 65 Resp: 19206
Ion Ratio Lower Upper
65 100
67 51.7 0.0 105.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.849 min Scan# 1142
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:114 Resp: 477257
Ion Ratio Lower Upper
114 100
63 21.2 0.0 40.2
88 16.2 0.0 32.0





#35
Dibromofluoromethane
Concen: 46.259 ug/l
RT: 7.904 min Scan# 987
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 113 Resp: 158775
Ion Ratio Lower Upper
113 100
111 104.7 81.8 122.8
192 17.6 13.0 19.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#36
1,1-Dichloropropene
Concen: 20.350 ug/l
RT: 8.093 min Scan# 1018
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 75 Resp: 94168
Ion Ratio Lower Upper
75 100
110 35.5 16.7 50.1
77 31.4 25.2 37.8





#37
Ethyl Acetate
Concen: 19.411 ug/l
RT: 7.270 min Scan# 884
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08





Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion: 43 Resp: 5957
Ion Ratio Lower Upper
43 100
61 13.8 10.5 15.7
70 10.8 8.5 12.7

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#38
Carbon Tetrachloride
Concen: 20.874 ug/l
RT: 8.081 min Scan# 1016
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:117 Resp: 90875
Ion Ratio Lower Upper
117 100
119 90.6 73.9 110.9
121 28.7 23.0 34.6





#39
Methylcyclohexane
Concen: 20.265 ug/l
RT: 9.343 min Scan# 11
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 83 Resp: 108509
Ion Ratio Lower Upper
83 100
55 84.3 65.0 97.6
98 48.7 40.8 61.2

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#40
Benzene
Concen: 20.954 ug/l
RT: 8.337 min Scan# 1058
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 78 Resp: 305208
Ion Ratio Lower Upper
78 100
77 23.4 18.7 28.1





#41
Methacrylonitrile
Concen: 18.101 ug/l
RT: 7.496 min Scan# 91
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08





Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion: 41	Resp: 31810
Ion Ratio	Lower Upper
41 100	
39 53.0	46.9 70.3
67 75.0	59.0 88.4
52 34.1	24.9 37.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#42
1,2-Dichloroethane
Concen: 20.142 ug/l
RT: 8.410 min Scan# 1070
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 62	Resp: 97519
Ion Ratio	Lower Upper
62 100	
98 9.5	0.0 18.4





#43
Isopropyl Acetate
Concen: 18.768 ug/l
RT: 8.435 min Scan# 1074
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 43 Resp: 103104
Ion Ratio Lower Upper
43 100
61 30.6 24.2 36.2
87 13.6 10.6 16.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#44
Trichloroethene
Concen: 20.915 ug/l
RT: 9.099 min Scan# 1183
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:130 Resp: 71206
Ion Ratio Lower Upper
130 100
95 101.4 0.0 215.2





#45
1,2-Dichloropropane
Concen: 20.449 ug/l
RT: 9.374 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 63 Resp: 7714
Ion Ratio Lower Upper
63 100
65 29.8 24.9 37.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#46
Dibromomethane
Concen: 19.993 ug/l
RT: 9.465 min Scan# 1243
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 93 Resp: 46372
Ion Ratio Lower Upper
93 100
95 83.5 66.6 99.8
174 83.5 69.0 103.6





#47
Bromodichloromethane
Concen: 20.093 ug/l
RT: 9.648 min Scan# 1273
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 83 Resp: 10629
Ion Ratio Lower Upper
83 100
85 62.6 50.3 75.5
127 7.5 5.6 8.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#48
Methyl methacrylate
Concen: 18.692 ug/l
RT: 9.441 min Scan# 1239
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 41 Resp: 48310
Ion Ratio Lower Upper
41 100
69 91.1 71.4 107.0
39 59.8 46.9 70.3





#49
 1,4-Dioxane
 Concen: 372.877 ug/l
 RT: 9.459 min Scan# 1242
 Delta R.T. -0.006 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01






Tgt Ion: 88 Resp: 1158
 Ion Ratio Lower Upper
 88 100
 43 29.6 0.0 0.0
 58 84.1 62.9 94.3

Manual Integrations
 APPROVED




Reviewed By :Mahesh
 Dadoda
 09/29/2025


Supervised By :Semsettin
 Yesilyurt
 09/29/2025



#50
 Toluene-d8
 Concen: 46.852 ug/l
 RT: 10.325 min Scan# 1384
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08


Tgt Ion: 98 Resp: 579559
 Ion Ratio Lower Upper
 98 100
 100 67.5 53.0 79.4





#51
4-Methyl-2-Pentanone
Concen: 91.920 ug/l
RT: 10.215 min Scan# 1366
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 43 Resp: 29940
Ion Ratio Lower Upper
43 100
58 39.3 32.3 48.5


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025



Supervised By :Semsettin
Yesilyurt
09/29/2025




#52
Toluene
Concen: 20.760 ug/l
RT: 10.392 min Scan# 1395
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08



Tgt Ion: 92 Resp: 187424
Ion Ratio Lower Upper
92 100
91 172.1 139.2 208.8





#53
t-1,3-Dichloropropene
Concen: 20.117 ug/l
RT: 10.611 min Scan# 1431
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08





Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion: 75 Resp: 9750
Ion Ratio Lower Upper
75 100
77 33.4 24.6 37.0

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#54
cis-1,3-Dichloropropene
Concen: 20.403 ug/l
RT: 10.075 min Scan# 1343
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 75 Resp: 114169
Ion Ratio Lower Upper
75 100
77 33.1 26.2 39.4
39 50.7 39.0 58.6





#55
1,1,2-Trichloroethane
Concen: 19.344 ug/l
RT: 10.788 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 97 Resp: 6124

Ion	Ratio	Lower	Upper
97	100		
83	91.5	67.4	101.2
85	59.2	42.0	63.0
99	62.7	49.0	73.6

Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025




Supervised By :Semsettin
Yesilyurt
09/29/2025




#56
Ethyl methacrylate
Concen: 19.046 ug/l
RT: 10.648 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 69 Resp: 79427


Ion	Ratio	Lower	Upper
69	100		
41	67.8	54.7	82.1
39	33.3	26.3	39.5





#57
1,3-Dichloropropane
Concen: 19.749 ug/l
RT: 10.934 min Scan# 1484
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01




Tgt Ion: 76 Resp: 10801
Ion Ratio Lower Upper
76 100
78 32.7 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025




#58
2-Chloroethyl Vinyl ether
Concen: 87.473 ug/l
RT: 9.928 min Scan# 1319
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Tgt Ion: 63 Resp: 200248
Ion Ratio Lower Upper
63 100
106 28.3 22.5 33.7





#59
2-Hexanone
Concen: 96.071 ug/l
RT: 10.971 min Scan# 1490
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01



Tgt Ion: 43 Resp: 21369
Ion Ratio Lower Upper
43 100
58 53.2 27.9 83.6




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025



#60
Dibromochloromethane
Concen: 20.044 ug/l
RT: 11.129 min Scan# 1516
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:129 Resp: 69693
Ion Ratio Lower Upper
129 100
127 76.9 39.3 117.8





#61
1,2-Dibromoethane
Concen: 19.678 ug/l
RT: 11.239 min Scan# 1534
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:107 Resp: 5902
Ion Ratio Lower Upper
107 100
109 95.8 74.1 111.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025



Supervised By :Semsettin
Yesilyurt
09/29/2025




#62
4-Bromofluorobenzene
Concen: 48.349 ug/l
RT: 12.617 min Scan# 1760
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Tgt Ion: 95 Resp: 225595
Ion Ratio Lower Upper
95 100
174 66.1 0.0 130.6
176 65.9 0.0 127.0





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.629 min Scan# 1159
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:117 Resp: 42668
Ion Ratio Lower Upper
117 100
82 56.6 46.9 70.3
119 30.9 24.9 37.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#64
Tetrachloroethene
Concen: 21.429 ug/l
RT: 10.861 min Scan# 1472
Delta R.T. -0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:164 Resp: 55720
Ion Ratio Lower Upper
164 100
166 133.6 96.9 145.3
129 108.6 75.0 112.4
131 100.2 67.4 101.0





#65
Chlorobenzene
Concen: 21.706 ug/l
RT: 11.660 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:112 Resp: 21333
Ion Ratio Lower Upper
112 100
114 30.5 24.8 37.2


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025



Supervised By :Semsettin
Yesilyurt
09/29/2025




#66
1,1,1,2-Tetrachloroethane
Concen: 20.847 ug/l
RT: 11.727 min Scan# 1614
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Tgt Ion:131 Resp: 64799
Ion Ratio Lower Upper
131 100
133 96.7 47.1 141.3
119 70.2 33.9 101.7





#67
Ethyl Benzene
Concen: 21.742 ug/l
RT: 11.727 min Scan# 1614
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01



Tgt Ion: 91 Resp: 349768
Ion Ratio Lower Upper
91 100
106 30.7 24.7 37.1




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025



#68
m/p-Xylenes
Concen: 44.189 ug/l
RT: 11.837 min Scan# 1632
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion:106 Resp: 272503
Ion Ratio Lower Upper
106 100
91 207.0 169.0 253.4









#69
o-Xylene
Concen: 21.555 ug/l
RT: 12.166 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion:106 Resp: 124923
Ion Ratio Lower Upper
106 100
91 221.5 107.8 323.6




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025



#70
Styrene
Concen: 22.230 ug/l
RT: 12.178 min Scan# 1688
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:104 Resp: 225931
Ion Ratio Lower Upper
104 100
78 51.3 40.8 61.2
103 54.6 43.4 65.2





#71
Bromoform
Concen: 19.735 ug/l
RT: 12.349 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:173 Resp: 35740
Ion Ratio Lower Upper
173 100
175 48.0 23.1 69.2
254 0.1 0.1 0.1


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025



Supervised By :Semsettin
Yesilyurt
09/29/2025




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.556 min Scan# 1914
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08




Tgt Ion:152 Resp: 223480
Ion Ratio Lower Upper
152 100
115 61.6 32.5 97.5
150 162.3 0.0 370.2





#73
Isopropylbenzene
Concen: 20.627 ug/l
RT: 12.464 min Scan# 1735
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 105 Resp: 32084
Ion Ratio Lower Upper
105 100
120 26.0 13.1 39.3

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#74
N-ethyl acetate
Concen: 19.100 ug/l
RT: 12.269 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 43 Resp: 93561
Ion Ratio Lower Upper
43 100
70 43.3 35.0 52.4
55 27.2 20.6 31.0
61 24.9 20.0 30.0





#75
 1,1,2,2-Tetrachloroethane
 Concen: 18.837 ug/l
 RT: 12.714 min Scan# 1776
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Instrument : MSVOA_W
 ClientSampleId : VW0926SBS01






Tgt Ion	Ratio	Lower	Upper
83	100		
131	10.1	4.9	14.7
85	64.7	31.9	95.7

Manual Integrations
 APPROVED




Reviewed By :Mahesh Dadoda
 09/29/2025


Supervised By :Semsettin Yesilyurt
 09/29/2025



#76
 1,2,3-Trichloropropane
 Concen: 18.325 ug/l m
 RT: 12.763 min Scan# 1784
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08


Tgt Ion	Ratio	Lower	Upper
75	100		
77	301.6	140.9	422.7





#77
Bromobenzene
Concen: 20.312 ug/l
RT: 12.745 min Scan# 1781
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:156 Resp: 77712
Ion Ratio Lower Upper
156 100
77 218.6 103.3 309.8
158 95.1 45.4 136.2

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#78
n-propylbenzene
Concen: 21.290 ug/l
RT: 12.800 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Tgt Ion: 91 Resp: 415955
Ion Ratio Lower Upper
91 100
120 21.8 10.9 32.9









#79
2-Chlorotoluene
Concen: 21.147 ug/l
RT: 12.891 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01

Tgt Ion: 91 Resp: 25028
Ion Ratio Lower Upper
91 100
126 32.5 16.3 48.8




Manual Integrations
APPROVED


Reviewed By :Mahesh
Dadoda
09/29/2025
Supervised By :Semsettin
Yesilyurt
09/29/2025



#80
1,3,5-Trimethylbenzene
Concen: 20.687 ug/l
RT: 12.940 min Scan# 1813
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:105 Resp: 275113
Ion Ratio Lower Upper
105 100
120 48.3 24.1 72.2





#81
trans-1,4-Dichloro-2-butene
Concen: 18.378 ug/l
RT: 12.513 min Scan# 11
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 75 Resp: 21754
Ion Ratio Lower Upper
75 100
53 112.1 81.4 122.2
89 43.9 35.4 53.2

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#82
4-Chlorotoluene
Concen: 20.849 ug/l
RT: 12.989 min Scan# 1821
Delta R.T. 0.006 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion: 91 Resp: 260792
Ion Ratio Lower Upper
91 100
126 33.1 16.5 49.5





#83
tert-Butylbenzene
Concen: 20.159 ug/l
RT: 13.202 min Scan# 1856
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01





Tgt Ion:119 Resp: 22420
Ion Ratio Lower Upper
119 100
91 72.1 35.5 106.5
134 25.7 13.5 40.4

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#84
1,2,4-Trimethylbenzene
Concen: 20.618 ug/l
RT: 13.245 min Scan# 1863
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:105 Resp: 278390
Ion Ratio Lower Upper
105 100
120 45.8 22.8 68.3





#85
 sec-Butylbenzene
 Concen: 20.833 ug/l
 RT: 13.379 min Scan# 1885
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0926SBS01






Tgt Ion:105 Resp: 347938
 Ion Ratio Lower Upper
 105 100
 134 19.2 9.8 29.5

Manual Integrations
 APPROVED




Reviewed By :Mahesh
 Dadoda
 09/29/2025


Supervised By :Semsettin
 Yesilyurt
 09/29/2025



#86
 p-Isopropyltoluene
 Concen: 20.974 ug/l
 RT: 13.495 min Scan# 1904
 Delta R.T. 0.000 min
 Lab File: VW032269.D
 Acq: 26 Sep 2025 10:08

Tgt Ion:119 Resp: 290111
 Ion Ratio Lower Upper
 119 100
 134 25.6 12.5 37.5
 91 24.4 12.1 36.3





#87

1,3-Dichlorobenzene

Concen: 21.606 ug/l

RT: 13.495 min Scan# 1904

Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Instrument :

MSVOA_W

ClientSampleId :

VW0926SBS01

Tgt Ion:146 Resp: 16350
 Ion Ratio Lower Upper
 146 100
 111 43.6 21.9 65.5
 148 61.8 31.9 95.9

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

09/29/2025

Supervised By :Semsettin

Yesilyurt

09/29/2025



#88

1,4-Dichlorobenzene

Concen: 20.581 ug/l




RT: 13.574 min Scan# 1917


Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08


Tgt Ion:146 Resp: 159112
 Ion Ratio Lower Upper
 146 100
 111 44.5 22.4 67.2
 148 65.8 34.4 103.1





#89
n-Butylbenzene
Concen: 20.855 ug/l
RT: 13.818 min Scan# 1957
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion: 91 Resp: 289490
Ion Ratio Lower Upper
91 100
92 52.4 26.7 80.1
134 23.8 12.3 36.8

Manual Integrations
APPROVED




Reviewed By :Mahesh
Dadoda
09/29/2025


Supervised By :Semsettin
Yesilyurt
09/29/2025



#90
Hexachloroethane
Concen: 20.770 ug/l
RT: 14.092 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08


Tgt Ion:117 Resp: 53545
Ion Ratio Lower Upper
117 100
201 60.8 31.4 94.0





#91
1,2-Dichlorobenzene
Concen: 20.903 ug/l
RT: 13.867 min Scan# 1965
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:146 Resp: 147974
Ion Ratio Lower Upper
146 100
111 43.7 21.9 65.8
148 60.9 31.9 95.5


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025



Supervised By :Semsettin
Yesilyurt
09/29/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 17.990 ug/l
RT: 14.476 min Scan# 2065
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08



Tgt Ion: 75 Resp: 12549
Ion Ratio Lower Upper
75 100
155 71.5 38.2 114.6
157 88.5 47.0 141.0





#93

1,2,4-Trichlorobenzene

Concen: 19.780 ug/l

RT: 15.123 min Scan# 2172

Delta R.T. -0.006 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08

Instrument :

MSVOA_W

ClientSampleId :

VW0926SBS01

Tgt Ion:180	Resp:	83072
Ion Ratio	Lower	Upper
180	100	
182	96.9	48.4 145.3
145	33.6	16.9 50.7

Manual Integrations

APPROVED

Reviewed By :Mahesh





Dadoda

09/29/2025

Supervised By :Semsettin

Yesilyurt

09/29/2025



#94

Hexachlorobutadiene

Concen: 21.539 ug/l




RT: 15.232 min Scan# 2189


Delta R.T. 0.000 min

Lab File: VW032269.D

Acq: 26 Sep 2025 10:08


Tgt Ion:225	Resp:	41545
Ion Ratio	Lower	Upper
225	100	
223	62.1	36.4 109.2
227	61.6	36.4 109.2





#95
Naphthalene
Concen: 18.725 ug/l
RT: 15.360 min Scan# 21
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08

Instrument :
MSVOA_W
ClientSampleId :
VW0926SBS01






Tgt Ion:128 Resp: 188749
Ion Ratio Lower Upper
128 100
127 12.7 10.6 15.8
129 11.3 9.0 13.4


Manual Integrations
APPROVED

Reviewed By :Mahesh
Dadoda
09/29/2025

Supervised By :Semsettin
Yesilyurt
09/29/2025



#96
1,2,3-Trichlorobenzene
Concen: 19.141 ug/l
RT: 15.543 min Scan# 2240
Delta R.T. 0.000 min
Lab File: VW032269.D
Acq: 26 Sep 2025 10:08



Tgt Ion:180 Resp: 74748
Ion Ratio Lower Upper
180 100
182 99.6 45.5 136.4
145 37.7 18.1 54.1

