

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092821\
 Data File : VW020215.D
 Acq On : 28 Sep 2021 15:03
 Operator : SY/VA
 Sample : M3834-14
 Misc : 5.05g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK002

Quant Time: Sep 29 00:59:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Sep 29 00:57:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.848	114	289853	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	279533	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	124773	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.361	65	118071	19.912	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.640%
7) Chloroethane-d5	2.898	69	95105	21.290	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.160%
11) 1,1-Dichloroethene-d2	4.025	63	140153	15.096	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.400%
21) 2-Butanone-d5	7.086	46	44806	48.966	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.940%
24) Chloroform-d	7.653	84	205140	22.504	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	8.311	65	117606	23.971	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.880%
32) Benzene-d6	8.281	84	371324	20.821	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.280%
36) 1,2-Dichloropropane-d6	9.281	67	116803	22.631	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.520%
41) Toluene-d8	10.329	98	331578	20.434	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	81.720%
43) trans-1,3-Dichloroprop...	10.579	79	47032	21.320	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.280%
47) 2-Hexanone-d5	10.927	63	33641	47.100	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.200%
56) 1,1,2,2-Tetrachloroeth...	12.695	84	95825	24.070	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.280%
66) 1,2-Dichlorobenzene-d4	13.853	152	115478	24.682	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	98.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092821\
Data File : VW020215.D
Acq On : 28 Sep 2021 15:03
Operator : SY/VA
Sample : M3834-14
Misc : 5.05g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK002

Quant Time: Sep 29 00:59:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092121SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Sep 29 00:57:09 2021
Response via : Initial Calibration

