

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
 Data File : VW024678.D
 Acq On : 29 Sep 2022 15:40
 Operator : SY/VA
 Sample : N4852-19
 Misc : 4.33g/10mL/MSVOA_W/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW822

Quant Time: Sep 30 02:38:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 30 02:36:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	539104	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	475033	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	204740	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	85034	15.076	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	60.320%		
7) Chloroethane-d5	2.885	69	52486	16.013	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	64.040%		
11) 1,1-Dichloroethene-d2	4.019	63	119607	9.162	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	36.640%#		
21) 2-Butanone-d5	7.080	46	120104	59.226	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	118.460%		
24) Chloroform-d	7.647	84	303821	21.081	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	84.320%		
26) 1,2-Dichloroethane-d4	8.305	65	167897	22.457	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	89.840%		
32) Benzene-d6	8.275	84	518799	17.087	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	68.360%		
36) 1,2-Dichloropropane-d6	9.275	67	197875	21.330	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	85.320%		
41) Toluene-d8	10.323	98	401671	14.802	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	59.200%		
43) trans-1,3-Dichloroprop...	10.579	79	75450	18.263	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	73.040%		
47) 2-Hexanone-d5	10.921	63	94311	59.418	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	118.840%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	189520	28.672	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	114.680%		
66) 1,2-Dichlorobenzene-d4	13.853	152	149697	20.778	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	83.120%		
Target Compounds						Qvalue
16) Methylene chloride	4.916	84	76535	7.229	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
Data File : VW024678.D
Acq On : 29 Sep 2022 15:40
Operator : SY/VA
Sample : N4852-19
Misc : 4.33g/10mL/MSVOA_W/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW822

Quant Time: Sep 30 02:38:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 30 02:36:56 2022
Response via : Initial Calibration

