

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
 Data File : VW024676.D
 Acq On : 29 Sep 2022 14:52
 Operator : SY/VA
 Sample : N4852-13
 Misc : 4.66g/10mL/MSVOA_W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW819

Quant Time: Sep 30 02:38:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 30 02:36:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	529405	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	460410	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	198842	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	93322	16.848	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.400%
7) Chloroethane-d5	2.891	69	60635	18.838	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.360%
11) 1,1-Dichloroethene-d2	4.019	63	131272	10.240	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	40.960%#
21) 2-Butanone-d5	7.080	46	122877	61.704	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	123.400%
24) Chloroform-d	7.653	84	325448	22.995	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.000%
26) 1,2-Dichloroethane-d4	8.305	65	179486	24.447	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	97.800%
32) Benzene-d6	8.275	84	541356	18.396	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.600%
36) 1,2-Dichloropropane-d6	9.274	67	206563	22.974	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.880%
41) Toluene-d8	10.323	98	412593	15.687	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	62.760%
43) trans-1,3-Dichloroprop...	10.579	79	78295	19.554	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.200%
47) 2-Hexanone-d5	10.920	63	96096	62.466	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	124.940%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	198293	30.952	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	123.800%#
66) 1,2-Dichlorobenzene-d4	13.853	152	151216	21.611	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
Data File : VW024676.D
Acq On : 29 Sep 2022 14:52
Operator : SY/VA
Sample : N4852-13
Misc : 4.66g/10mL/MSVOA_W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW819

Quant Time: Sep 30 02:38:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 30 02:36:56 2022
Response via : Initial Calibration

