

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
 Data File : VW024684.D
 Acq On : 29 Sep 2022 18:05
 Operator : SY/VA
 Sample : VIBLK559
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK559

Quant Time: Sep 30 02:40:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 30 02:36:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	546391	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	506611	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	244576	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	141501	24.752	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	99.000%
7) Chloroethane-d5	2.885	69	82091	24.711	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.840%
11) 1,1-Dichloroethene-d2	4.019	63	217115	16.409	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.640%
21) 2-Butanone-d5	7.080	46	105258	51.213	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.420%
24) Chloroform-d	7.653	84	330612	22.634	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.520%
26) 1,2-Dichloroethane-d4	8.305	65	178386	23.542	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.160%
32) Benzene-d6	8.275	84	714262	22.058	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.240%
36) 1,2-Dichloropropane-d6	9.275	67	218278	22.063	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.240%
41) Toluene-d8	10.323	98	623605	21.548	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	10.579	79	92390	20.970	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.880%
47) 2-Hexanone-d5	10.921	63	82500	48.737	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.480%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	171234	24.291	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	97.160%
66) 1,2-Dichlorobenzene-d4	13.847	152	189684	22.040	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	88.160%
Target Compounds						Qvalue
16) Methylene chloride	4.928	84	54558	5.084	ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092922\
Data File : VW024684.D
Acq On : 29 Sep 2022 18:05
Operator : SY/VA
Sample : VIBLK559
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK559


Quant Time: Sep 30 02:40:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092622SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 30 02:36:56 2022
Response via : Initial Calibration





#16
Methylene chloride
Concen: 5.084 ug/L
RT: 4.928 min Scan# 49
Delta R.T. 0.006 min
Lab File: VW024684.D
Acq: 29 Sep 2022 18:05

Instrument :
MSVOA_W
ClientSampleId :
VIBLK559



Tgt Ion: 84 Resp: 54558

Ion	Ratio	Lower	Upper
84	100		
86	58.7	44.0	81.8
49	118.0	87.4	162.2

