

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
 Data File : VW032289.D
 Acq On : 29 Sep 2025 10:18
 Operator : SY/MD
 Sample : VW0929SBL01
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VW0929SBL01

Quant Time: Sep 30 02:34:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.953	168	170986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	366797	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	375398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	185238	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	169881	62.923	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	125.840%
35) Dibromofluoromethane	7.898	113	134272	50.901	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	101.800%
50) Toluene-d8	10.325	98	493728	51.934	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	103.860%
62) 4-Bromofluorobenzene	12.617	95	177961	49.626	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	99.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
Data File : VW032289.D
Acq On : 29 Sep 2025 10:18
Operator : SY/MD
Sample : VW0929SBL01
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VW0929SBL01

Quant Time: Sep 30 02:34:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

