

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
 Data File : VW032292.D
 Acq On : 29 Sep 2025 11:52
 Operator : SY/MD
 Sample : Q3221-01RE
 Misc : 5.22g/5mL/MSVOA_W/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 HD-01-092625RE

Quant Time: Sep 30 02:36:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.953	168	69334	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	137638	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	131382	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.562	152	55223	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.313	65	59140	54.021	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	108.040%
35) Dibromofluoromethane	7.898	113	50003	50.516	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	101.040%
50) Toluene-d8	10.325	98	164485	46.108	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	92.220%
62) 4-Bromofluorobenzene	12.617	95	53876	40.038	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	80.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
Data File : VW032292.D
Acq On : 29 Sep 2025 11:52
Operator : SY/MD
Sample : Q3221-01RE
Misc : 5.22g/5mL/MSVOA_W/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
HD-01-092625RE

Quant Time: Sep 30 02:36:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

