

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
 Data File : VW032295.D
 Acq On : 29 Sep 2025 14:38
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.00g/5mL/MSVOA_W/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK

Quant Time: Sep 30 02:37:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	176727	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	398693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	413063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	191812	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	168278	60.304	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	120.600%
35) Dibromofluoromethane	7.904	113	141273	49.271	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	98.540%
50) Toluene-d8	10.331	98	516632	49.995	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	100.000%
62) 4-Bromofluorobenzene	12.617	95	179600	46.076	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	92.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW092925\
Data File : VW032295.D
Acq On : 29 Sep 2025 14:38
Operator : SY/MD
Sample : VIBLK
Misc : 5.00g/5mL/MSVOA_W/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK

Quant Time: Sep 30 02:37:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

