

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
 Data File : VW030290.D
 Acq On : 30 Sep 2024 12:38
 Operator : SY/MD
 Sample : VW0930SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK559

Quant Time: Oct 01 01:02:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	151828	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	142626	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	66146	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	33000	27.815	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 111.240%		
7) Chloroethane-d5	2.905	69	23177	26.849	ug/L	0.01
Spiked Amount	25.000	Range 30 - 150	Recovery	= 107.400%		
11) 1,1-Dichloroethene-d2	4.021	65	22688	26.197	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 104.800%		
21) 2-Butanone-d5	7.081	46	23336	48.910	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 97.820%		
24) Chloroform-d	7.654	84	100148	24.553	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 98.200%		
26) 1,2-Dichloroethane-d4	8.307	65	58158	27.317	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 109.280%		
32) Benzene-d6	8.276	84	195521	24.459	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 97.840%		
36) 1,2-Dichloropropane-d6	9.276	67	60263	25.065	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 100.280%		
41) Toluene-d8	10.325	98	178546	24.358	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 97.440%		
43) trans-1,3-Dichloroprop...	10.581	79	25441	25.440	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 101.760%		
47) 2-Hexanone-d5	10.922	63	19211	48.389	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 96.780%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	45001	25.048	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 100.200%		
66) 1,2-Dichlorobenzene-d4	13.848	152	58655	26.708	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 106.840%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
Data File : VW030290.D
Acq On : 30 Sep 2024 12:38
Operator : SY/MD
Sample : VW0930SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK559

Quant Time: Oct 01 01:02:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

