

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
 Data File : VW030294.D
 Acq On : 30 Sep 2024 15:33
 Operator : SY/MD
 Sample : P4183-20
 Misc : 5.24g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EM8

Quant Time: Oct 01 01:03:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	228954	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	174341	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	51148	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	53289	29.785	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 119.160%		
7) Chloroethane-d5	2.899	69	36789	28.261	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 113.040%		
11) 1,1-Dichloroethene-d2	4.027	65	34802	26.648	ug/L	0.01
Spiked Amount	25.000	Range 45 - 110	Recovery	= 106.600%		
21) 2-Butanone-d5	7.081	46	23148	32.173	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 64.340%		
24) Chloroform-d	7.648	84	142924	23.236	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 92.960%		
26) 1,2-Dichloroethane-d4	8.307	65	78832	24.554	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 98.200%		
32) Benzene-d6	8.276	84	270777	27.711	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 110.840%		
36) 1,2-Dichloropropane-d6	9.276	67	83252	28.328	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 113.320%		
41) Toluene-d8	10.325	98	231401	25.826	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 103.320%		
43) trans-1,3-Dichloroprop...	10.581	79	28335	23.180	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 92.720%		
47) 2-Hexanone-d5	10.922	63	18886	38.917	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 77.840%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	46661	21.248	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 85.000%		
66) 1,2-Dichlorobenzene-d4	13.849	152	37287	21.957	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 87.840%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
Data File : VW030294.D
Acq On : 30 Sep 2024 15:33
Operator : SY/MD
Sample : P4183-20
Misc : 5.24g/10mL/MSVOA_W/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EM8

Quant Time: Oct 01 01:03:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

