

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
 Data File : VW030295.D
 Acq On : 30 Sep 2024 15:57
 Operator : SY/MD
 Sample : P4183-21
 Misc : 5.65g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EM9

Quant Time: Oct 01 01:03:23 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	253953	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	223134	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	97155	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	54606	27.517	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	110.080%	
7) Chloroethane-d5	2.899	69	37520	25.986	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	103.960%	
11) 1,1-Dichloroethene-d2	4.021	65	36549	25.231	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	100.920%	
21) 2-Butanone-d5	7.075	46	34551	43.295	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	86.580%	
24) Chloroform-d	7.648	84	158343	23.209	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	92.840%	
26) 1,2-Dichloroethane-d4	8.307	65	88795	24.935	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	99.760%	
32) Benzene-d6	8.276	84	299348	23.936	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	95.760%	
36) 1,2-Dichloropropane-d6	9.270	67	92863	24.689	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	98.760%	
41) Toluene-d8	10.325	98	268821	23.442	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	93.760%	
43) trans-1,3-Dichloroprop...	10.575	79	38050	24.321	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	97.280%	
47) 2-Hexanone-d5	10.922	63	27162	43.731	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	87.460%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	63393	22.554	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	90.200%	
66) 1,2-Dichlorobenzene-d4	13.849	152	76344	23.668	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	94.680%	
Target Compounds					Qvalue	
13) Acetone	4.125	43	4111	4.856	ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093024\
Data File : VW030295.D
Acq On : 30 Sep 2024 15:57
Operator : SY/MD
Sample : P4183-21
Misc : 5.65g/10mL/MSVOA_W/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EM9

Quant Time: Oct 01 01:03:23 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

