

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093025\
 Data File : VW032303.D
 Acq On : 30 Sep 2025 15:14
 Operator : SY/MD
 Sample : Q3150-04
 Misc : 5.00g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 Mid Site Grid 5-7

Quant Time: Oct 01 00:54:26 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	73131	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.849	114	148995	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.635	117	140500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	58965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	60653	52.526	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	105.060%
35) Dibromofluoromethane	7.904	113	49577	46.268	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	92.540%
50) Toluene-d8	10.324	98	166304	43.064	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	86.120%
62) 4-Bromofluorobenzene	12.617	95	54149	37.173	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	74.340%
Target Compounds						
					Qvalue	
16) Acetone	4.167	43	4357	14.692	ug/l	92
20) Methylene Chloride	4.947	84	18998	11.998	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093025\
Data File : VW032303.D
Acq On : 30 Sep 2025 15:14
Operator : SY/MD
Sample : Q3150-04
Misc : 5.00g/5mL/MSVOA_W/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
Mid Site Grid 5-7

Quant Time: Oct 01 00:54:26 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

