

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093025\
 Data File : VW032305.D
 Acq On : 30 Sep 2025 15:57
 Operator : SY/MD
 Sample : Q3245-02
 Misc : 3.98g/5mL/MSVOA_W/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 13-1

Quant Time: Oct 01 00:55:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 26 07:11:42 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.965	168	184640	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	416613	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.629	117	444982	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.556	152	232264	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.319	65	175974	60.360	ug/l	0.00
Spiked Amount	50.000	Range	63 - 155	Recovery	=	120.720%
35) Dibromofluoromethane	7.904	113	144646	48.277	ug/l	0.00
Spiked Amount	50.000	Range	70 - 134	Recovery	=	96.560%
50) Toluene-d8	10.325	98	563467	52.182	ug/l	0.00
Spiked Amount	50.000	Range	74 - 123	Recovery	=	104.360%
62) 4-Bromofluorobenzene	12.617	95	240360	59.012	ug/l	0.00
Spiked Amount	50.000	Range	17 - 146	Recovery	=	118.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW093025\
Data File : VW032305.D
Acq On : 30 Sep 2025 15:57
Operator : SY/MD
Sample : Q3245-02
Misc : 3.98g/5mL/MSVOA_W/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
13-1

Quant Time: Oct 01 00:55:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\82W092525S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 26 07:11:42 2025
Response via : Initial Calibration

