

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
 Data File : VW030309.D
 Acq On : 01 Oct 2024 14:30
 Operator : SY/MD
 Sample : P4183-08
 Misc : 5.00g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 02 07:42:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	158270	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	153745	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	69296	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	35444	28.659	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	114.640%
7) Chloroethane-d5	2.899	69	25876	28.755	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	115.040%
11) 1,1-Dichloroethene-d2	4.021	65	22636	25.074	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	100.280%
21) 2-Butanone-d5	7.081	46	26223	52.724	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.440%
24) Chloroform-d	7.648	84	106548	25.059	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.240%
26) 1,2-Dichloroethane-d4	8.307	65	61016	27.493	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.960%
32) Benzene-d6	8.276	84	204556	23.739	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.960%
36) 1,2-Dichloropropane-d6	9.270	67	64869	25.030	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	100.120%
41) Toluene-d8	10.325	98	185684	23.500	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.000%
43) trans-1,3-Dichloroprop...	10.581	79	24678	22.893	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	91.560%
47) 2-Hexanone-d5	10.922	63	20713	48.399	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.800%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	46852	24.193	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.760%
66) 1,2-Dichlorobenzene-d4	13.848	152	59364	25.803	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
Data File : VW030309.D
Acq On : 01 Oct 2024 14:30
Operator : SY/MD
Sample : P4183-08
Misc : 5.00g/10mL/MSVOA_W/SOIL/A
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Oct 02 07:42:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

