

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
 Data File : VW030302.D
 Acq On : 01 Oct 2024 10:06
 Operator : SY/MD
 Sample : VW1001SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK560

Quant Time: Oct 02 07:42:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	159446	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	149725	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	69678	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	37129	29.800	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 119.200%		
7) Chloroethane-d5	2.899	69	28384	31.310	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 125.240%		
11) 1,1-Dichloroethene-d2	4.021	65	23584	25.931	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 103.720%		
21) 2-Butanone-d5	7.081	46	23518	46.937	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 93.880%		
24) Chloroform-d	7.654	84	108963	25.438	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 101.760%		
26) 1,2-Dichloroethane-d4	8.307	65	59248	26.499	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 106.000%		
32) Benzene-d6	8.276	84	209858	25.008	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 100.040%		
36) 1,2-Dichloropropane-d6	9.276	67	64987	25.748	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 103.000%		
41) Toluene-d8	10.325	98	190880	24.806	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 99.240%		
43) trans-1,3-Dichloroprop...	10.581	79	24694	23.522	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 94.080%		
47) 2-Hexanone-d5	10.922	63	18245	43.777	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 87.560%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	42803	22.695	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 90.800%		
66) 1,2-Dichlorobenzene-d4	13.848	152	59431	25.690	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 102.760%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
Data File : VW030302.D
Acq On : 01 Oct 2024 10:06
Operator : SY/MD
Sample : VW1001SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK560

Quant Time: Oct 02 07:42:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

