

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
 Data File : VW030305.D
 Acq On : 01 Oct 2024 12:18
 Operator : SY/MD
 Sample : P4183-22RE
 Misc : 4.86g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EN2RE

Quant Time: Oct 02 07:42:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	210794	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	140070	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	34477	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	56990	34.598	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	= 138.400%		
7) Chloroethane-d5	2.905	69	38675	32.270	ug/L	0.01
Spiked Amount	25.000	Range 30 - 150	Recovery	= 129.080%		
11) 1,1-Dichloroethene-d2	4.021	65	32009	26.621	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	= 106.480%		
21) 2-Butanone-d5	7.081	46	27921	42.150	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 84.300%		
24) Chloroform-d	7.654	84	129994	22.955	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	= 91.840%		
26) 1,2-Dichloroethane-d4	8.307	65	74537	25.217	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	= 100.880%		
32) Benzene-d6	8.276	84	247033	31.467	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	= 125.880%		
36) 1,2-Dichloropropane-d6	9.276	67	76393	32.354	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	= 129.400%#		
41) Toluene-d8	10.325	98	189536	26.329	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	= 105.320%		
43) trans-1,3-Dichloroprop...	10.581	79	27755	28.261	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	= 113.040%		
47) 2-Hexanone-d5	10.922	63	22104	56.692	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	= 113.380%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	40209	22.790	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	= 91.160%		
66) 1,2-Dichlorobenzene-d4	13.849	152	25218	22.031	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	= 88.120%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
Data File : VW030305.D
Acq On : 01 Oct 2024 12:18
Operator : SY/MD
Sample : P4183-22RE
Misc : 4.86g/10mL/MSVOA_W/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EN2RE

Quant Time: Oct 02 07:42:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

