

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
 Data File : VW030315.D
 Acq On : 01 Oct 2024 16:56
 Operator : SY/MD
 Sample : P4237-10
 Misc : 4.91g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4F24

Quant Time: Oct 02 07:43:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	246096	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	169424	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	45933	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	59043	30.703	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	122.800%
7) Chloroethane-d5	2.899	69	42451	30.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	121.360%
11) 1,1-Dichloroethene-d2	4.027	65	35859	25.545	ug/L	0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	102.200%
21) 2-Butanone-d5	7.081	46	23933	30.947	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	61.900%
24) Chloroform-d	7.654	84	145142	21.953	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.800%
26) 1,2-Dichloroethane-d4	8.307	65	84751	24.559	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.240%
32) Benzene-d6	8.276	84	275693	29.033	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	116.120%
36) 1,2-Dichloropropane-d6	9.276	67	88357	30.937	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	123.760%#
41) Toluene-d8	10.325	98	217730	25.006	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.040%
43) trans-1,3-Dichloroprop...	10.581	79	27001	22.730	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.920%
47) 2-Hexanone-d5	10.922	63	18440	39.100	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.200%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	48113	22.545	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.160%
66) 1,2-Dichlorobenzene-d4	13.854	152	32489	21.304	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
Data File : VW030315.D
Acq On : 01 Oct 2024 16:56
Operator : SY/MD
Sample : P4237-10
Misc : 4.91g/10mL/MSVOA_W/SOIL/A
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4F24

Quant Time: Oct 02 07:43:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

