

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
 Data File : VW030318.D
 Acq On : 01 Oct 2024 18:08
 Operator : SY/MD
 Sample : P4237-14
 Misc : 4.89g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 A4EN6

Quant Time: Oct 02 07:43:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Oct 01 01:01:52 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.843	114	268251	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.629	117	180561	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.556	152	47232	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.363	65	58975	28.134	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.520%	
7) Chloroethane-d5		2.905	69	41652	27.310	ug/L	0.01
Spiked Amount	25.000	Range	30 - 150	Recovery	=	109.240%	
11) 1,1-Dichloroethene-d2		4.021	65	34545	22.577	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	90.320%	
21) 2-Butanone-d5		7.075	46	23120	27.427	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	54.860%	
24) Chloroform-d		7.648	84	144828	20.097	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	80.400%	
26) 1,2-Dichloroethane-d4		8.307	65	78077	20.757	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.040%	
32) Benzene-d6		8.276	84	271975	26.875	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	107.520%	
36) 1,2-Dichloropropane-d6		9.276	67	85013	27.931	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	111.720%	
41) Toluene-d8		10.325	98	218177	23.511	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	94.040%	
43) trans-1,3-Dichloroprop...		10.575	79	23643	18.675	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	74.720%	
47) 2-Hexanone-d5		10.922	63	17236	34.293	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	68.580%	
56) 1,1,2,2-Tetrachloroeth...		12.690	84	42523	18.696	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	74.800%	
66) 1,2-Dichlorobenzene-d4		13.848	152	28799	18.365	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	73.440%#	
Target Compounds							Qvalue
13) Acetone		4.125	43	3387	3.788	ug/L	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100124\
Data File : VW030318.D
Acq On : 01 Oct 2024 18:08
Operator : SY/MD
Sample : P4237-14
Misc : 4.89g/10mL/MSVOA_W/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
A4EN6

Quant Time: Oct 02 07:43:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Oct 01 01:01:52 2024
Response via : Initial Calibration

