

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
 Data File : VW030325.D
 Acq On : 02 Oct 2024 14:25
 Operator : SY/MD
 Sample : VW1002SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Oct 03 04:11:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Oct 03 04:10:12 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	183215	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	170493	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	73939	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	40129	28.029	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	112.120%
7) Chloroethane-d5	2.899	69	31638	30.372	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	121.480%
11) 1,1-Dichloroethene-d2	4.027	65	25061	23.980	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	95.920%
21) 2-Butanone-d5	7.081	46	26424	45.895	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.780%
24) Chloroform-d	7.654	84	114160	23.193	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.760%
26) 1,2-Dichloroethane-d4	8.307	65	62935	24.497	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	98.000%
32) Benzene-d6	8.276	84	224152	23.458	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	93.840%
36) 1,2-Dichloropropane-d6	9.276	67	71208	24.777	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.120%
41) Toluene-d8	10.325	98	204920	23.387	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.560%
43) trans-1,3-Dichloroprop...	10.581	79	26489	22.159	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.640%
47) 2-Hexanone-d5	10.922	63	20580	43.364	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	48457	22.564	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	90.240%
66) 1,2-Dichlorobenzene-d4	13.854	152	63195	25.743	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW100224\
Data File : VW030325.D
Acq On : 02 Oct 2024 14:25
Operator : SY/MD
Sample : VW1002SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Oct 03 04:11:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM092724SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Oct 03 04:10:12 2024
Response via : Initial Calibration

